

Invitation of quotation
for
General Hospital Equipments
At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/29-05/2017-AIIMS.JDH

Inquiry Issue Date : 03rd February, 2018

Last Date of Submission : 09th February, 2018 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

**Invitation of quotation for General Hospital Equipments at
AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for General Hospital Equipments for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 09.02.2018 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR STEAM CHAMBER SITTING AND DHARA
STAND FOR AYUSHAGAINST INQUIRY NO. ADMN/GEN/29-05/2017-
AIIMS.JDH” DUE ON 09.02.2018 05.00 PM”**

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/VAT/CST/ST/Other taxes and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**

INQUIRY NO. Admin/Gen/29-04/2017-AIIMS.JDH

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**
- D) **Inspection committee will check the products thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then AIIMS, Jodhpur has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1
Specification

S. No.	Particular	Specification	Qty.
1	Peak Flow Meter for Spirometry with FEVI	<ul style="list-style-type: none"> • Measuring Range: PEF from 50 to 900 l/min. FEV 1 from 0.01 to 9.99 • Measuring method: Rotating wing wheel • Accuracy: TAS 1994 updated ± 25 l/min or 12% of reading • PEF: 1 l/min, FEV 1: 0.01 • Memory: 240 measurements with date/time 	04 Unit
2	Nebulizer	<ul style="list-style-type: none"> • Power Consumption-1 A • Nebulizer Rate- 0.4 ml/min • Particle Size- 2-6 μm • Medication cup capacity • Appropriate Medication Quantities- 2-6 ml • Operating Temperature/Humidity- 10-40 °C, 30-85% RH • Dimension Approx. 158 (W) x 175(D) x 100(H) mm (Compressor only) • Usable to both adults and child • MMD approximately 5 micrometer • Contains: Compressor, nebulizer Kit, air tubes, 5 pieces replacement filters, mouth piece, adult mask, child mask, instruction manual and carrying bag • year manufacturer's warranty form the date of purchase 	04 Unit
3	Pulse Oximeter	<ul style="list-style-type: none"> • Should have plethismographic wave form with numeric display for SPO2 and Heart rate on LCD/TFT display. • Should have a SPO2 range of 0 to 100% • Should have SPO2 accuracy of $\pm 2\%$ • Should provide bar graph for pulse strength • Audio and visual alarm for both upper and lower SPO2 & Heart rate • Should provide with adult reusable finger probe with technology form standard reputed companies • Should be a portable & light weighted • Should work with input 200 to 240 Vac 50 Hz supply • Patient Range- Adult, Pediatrics and neonatal patients • Should have safety certificate form a competent authority CE/FDA (US)/ STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report form ERTL. Copy of the certificate/test report shall be product along with the technical bid. • 5 year manufacturer's warranty 	04 Unit

INQUIRY NO. Admin/Gen/29-04/2017-AIIMS.JDH

4	Mini Ophthalmoscope	<ul style="list-style-type: none"> • Illumination: 3.5V, 2.8W Mini Halogen bulb • Recharging unit: Input Voltage: 220V \pm 10% V • Input Frequency: 50 Hz \pm 1Hz • Input Power: 8VA • Viewing Lenses: 0, \pm1, \pm2,\pm3, \pm4, \pm5, \pm6, \pm8, \pm10, \pm12, \pm16, \pm20, -25, -35 • Apertures: Large Spot, Small Spot, Slit, Central Net and Red-free New • Weight: 118g (excluding battery) • Total Weight: 340g • Dimensions: 42mm x 32mm x 210mm\ 	04 Unit
5	2.5 V Basic Student Pocket Scope, Otoscope, Auriscope with AA Battery Handle	<ul style="list-style-type: none"> • 2.5v Halogen light Otoscope with Plastic Head and Handle • Movable Optical glass viewing window with 3x magnification • Plastic reusable aural speculum size 2 3, 4, 5, 6mm • AA size Battery Handle 	04 Unit
6	Infrared Thermometer	<ul style="list-style-type: none"> • Range: -60 to 500°C • Resolution: 0.1°C or °F (-9.9 to 199.9°C) or 1°C • Infrared Accuracy: \pm°C (0 to 65°C) or \pm2°C • Field of view: Target ratio 11:1 • Battery: 2 x AAA • Battery Life: 140 hours continuous use • Display: custom LCD • Dimension: 39 x 72 x 175mm • Weight: 180gms 	04 Unit
7	Electronic Baby Weighing Scale	<ul style="list-style-type: none"> • Manufactured by Medicare Products Inc • Electronic Digital Baby Infant weighing Scale • Smooth and curved platform for baby • Weight can be displayed in Kilograms (kg.) and ponds (lb) • The display will stabilize and flash • Display weight backlit LCD screen 	04 Unit
8	Infantometer	<ul style="list-style-type: none"> • Height measuring scale • Dual Scale-cm and inches • Size: 18" x 7" (Length x width) • Item weight: 2.44 Kg. • Broad acrylic base with one sliding side adjustable as per length of baby • Dual scale for direct reading in cms/inches upto 90cm • Folding sides present for accurate measurement of the length of the child 	04 Unit

INQUIRY NO. Admin/Gen/29-04/2017-AIIMS.JDH

9	Snellen's Chart	<ul style="list-style-type: none"> • Manufactured by BEXCO Snellen Eye chart Snellen chart • These opt types feature strong serifs and follow idea of the 5 x 5 grid • Snellen chart is also coupled with red and green stripes for an easy color vision test • Nonreflecting, matte finish eye charts, designed for vision testing at 10 feet distance • Eye chart measures 9 x 14 in • Plastic washable sheet • Visual acuity 20/100 to 20/6 	04 Unit
9	Heine Mini 3000 Otoscope with Plastic Handle with Disposable Tips (Single)	<ul style="list-style-type: none"> • Viewing window with 3x magnification and optimized casing surface for razor-shape images and minimum reflection. Swiveling viewing window built into instrument • for Examination of Ear • 2-piece handle and head, Easy to maintain, Flexible 	04 Unit
10	Personal Sampler with Cyclone attachment Envirotech APM801 (With 4 Packet of Whatman GF/A Grade Glass Microfiber filters 37 mm dia in a sealed packet of 100 discs.)	<p>Capable of sampling respirable dust (cut off size 5 microns) total dusts and metal fumes in work area or shop-floor environment and consisting of:</p> <p>Cyclone attachment for Respirable Dust, Sampling pump with rechargeable batteries, sampling head to accommodate 37mm dia filters a battery charger belt and a carrying case.</p> <ul style="list-style-type: none"> • Flow Rate: 0.5 – 3.0 L.PM. • Cyclone: Brass with Teflon filter holder to cut-off at 5 microns • Filter: 37 mm dia glass fiber filter discs • Operation Time: 8 hours on fully charged batteries • Battery: Ni-Cd rechargeable • Recharge Time: 12-14 hours after discharge • Size: 155 x 82 x 60mm • Weight 1.3kg approx. (with cyclone) 	01 Unit

Note: The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR GENERAL HOSPITAL EQUIPMENTS AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/29-05/2017-AIIMS.JDH" DUE ON 09.02.2018 at 05.00 PM for General Hospital Equipments at AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Peak Flow Meter for Spirometry with FEVI	04 Unit						
2.	Nebulizer	04 Unit						
3.	Pulse Oximeter	04 Unit						
4.	Mini Ophthalmoscope	04 Unit						
5.	2.5 V Basic Student Pocket Scope, Otoscope, Auriscope with AA Battery Handle	04 Unit						
6.	Infrared Thermometer	04 Unit						
7.	Electronic Baby Weighing Scale	04 Unit						
8.	Infantometer	04 Unit						
9.	Snellen's Chart	04 Unit						
10.	Heine Mini 3000 Otoscope with Plastic Handle with Disposable Tips (Single)	04 Unit						
11.	Personal Sampler with Cyclone attachment Envirotech APM801 (With 4 Packet of Whatman GF/A Grade Glass Microfiber filters 37 mm dia in a sealed packet of 100 discs.)	01 Unit						

Note:

1. The Bidder must quote only single Make & Model.
2. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

INQUIRY NO. Admin/Gen/29-04/2017-AIIMS.JDH

Date_____

Place_____

(Signature of Authorized Person)_____

(Name)_____

Name of Firm/Company/Agency_____

Phone No._____

Email:_____