

Invitation of quotation

for

PTCA Kit

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No. : Admin/Gen/12-03/2017-AIIMS.JDH

Inquiry Issue Date : 12th April, 2017

Last Date of Submission : 18th April, 2017 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for PTCA Kit at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for supply of PTCA Kit for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 18.04.2017 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR SUPPLY OF PTCA KIT AGAINST INQUIRY NO. ADMN/GEN/12-03/2017-AIIMS.JDH” DUE ON 18.04.2017 05.00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid VAT/ Sales Tax No. and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior

INQUIRY NO. Admin/Gen/12-03/2017-AIIMS.JDH

confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The bidder must be Authorized Distributor/Dealer. (Certificate must be attached by Bidder).**
- D) **The supplier may be asked to submit a sample of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1
Specification

Sr. No	Description Of Item	Recommended Make	Specification	Qty
1	PCTA Kit Containing- (i) Three Port manifold With Knobs to turn "Right" When Open (ii) One Pressure Line (iii) Fluid Connecting Line (iv) Contrast Connecting Line (v) One Three Way Stopcock (vi) Short Tube of 12 to 15 Inches to Connect to Touhy Borst (vii) One Y-Connector Hemostatic Valve with Spring Type Push and Release Mechanism (viii) One Inflation Device with Manometer up to 30 Atmosphere (Easy to Operate with Luminescent Dial) (ix) One Luer Lock Controlled Syringe of 10 ml with Finger Grip (x) Insertion Needle (xi) Torque Device	All Companies (Inflation Device should be from Abbott, Cordis, Boston and Medtronics Only)	--	20 Nos
2.	Angiographic Guide wire, PTFE Coated, 'J' tip, Diameter 0.032"	Medtronic, Terumo, Abbott, Boston, Asahi	US FDA Approved Size: 150 cm	30 Nos

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

I/We Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF PTCA KIT AGAINST THE INQUIRY NO. Admn/Gen/12-03/2017-AIIMS.JDH" DUE ON 18.04.2017 05.00 PM for supply of PTCA Kit at AIIMS Jodhpur".

1. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
2. I/We hereby offer to supply at the following rates.

S. No.	Particulars	Quoted Make	Quoted Pack Size	Qty	Price/Unit Exclusive of TAX (INR)	TAX %	Price/Unit inclusive of TAX (INR)	Total Amount Inclusive of TAX (INR)
1	PCTA Kit Containing- (i) Three Port manifold With Knobs to turn "Right" When Open (ii) One Pressure Line (iii) Fluid Connecting Line (iv) Contrast Connecting Line (v) One Three Way Stopcock (vi) Short Tube of 12 to 15 Inches to Connect to Touhy Borst (vii) One Y-Connector Hemostatic Valve with Spring Type Push and Release Mechanism (viii) One Inflation Device with Manometer up to 30 Atmosphere (Easy to Operate with Luminescent Dial) (ix) One Luer Lock Controlled Syringe of 10 ml with Finger Grip (x) Insertion Needle (xi) Torque Device			20 Nos				
2.	Angiographic Guide wire, PTFE Coated, 'J' tip, Diameter 0.032"			30 Nos				

Note: L1 will be decided on composite basis / L1 will be decided on individual item wise.

Date _____

Place _____

(Signature of Authorized Person) _____

(Name) _____

Name of Firm/Company/Agency _____

Phone No. _____

Email: _____