

Invitation of quotation
for
Kits for Immunopathology Test
At

All India Institute of Medical Sciences, Jodhpur

Inquiry No. : Admin/Gen/36-02/2020-AIIMS.JDH

Inquiry Issue Date : 14th July, 2020

Last Date of Submission : 20th July, 2020 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

**Invitation of quotation for Kits for Immunopathology Test at
AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for supply of Kits for Immunopathology Test for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 20.07.2020 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“Quotation for supply of Kits for Immunopathology Test
AGAINST INQUIRY NO. ADMN/GEN/36-02/2020-AIIMS.JDH” DUE ON
20.07.2020 05.00 PM”**

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST No. and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**

INQUIRY NO. Admin/Gen/36-02/2020-AIIMS.JDH

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
B) **Catalog may be attached with quotation for technical evaluation.**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1
Specification

S.No	Test	Kit Pack Size
1	ANA IIF	10X05
2	ANA immunoblot	16x1
3	ANCA IIF	10X05
4	PLA2R IIF	10X03
5	Mosaic basic profile	10X5
6	IIF dsDNA	10X5
7	MPO/PR3/GBM profile	16 strip

Test specifications for Immunopathology tests

ANA-IFA kit: 50 test/kit

1. Should have Hep-20-10 cells
2. Should have much higher proportion of mitotic cells.
3. Should have facility to upgrade the kit with multiple substrates.
4. Should have Primate liver as additional substrate for parallel detection and confirmation of ANA and liver antibodies.
5. Should have reagent incubation tray for test incubation.
6. Test system should be based on the titerplane technique and should not have requirement of humidity chamber.
7. Should be stable for 12-18 months after manufacturing when stored at temperature between +2°C and +8°C.

ANCA-IFA kit: 50 test/kit

1. Should have ethanol fixed granulocytes and formalin fixed granulocytes in one reaction field.
2. Should have facility to upgrade the assay with multiple substrates.
3. Should have reagent tray system for assay incubation.
4. Test system should be based on the titerplane technique and be free from need of humidity chamber
5. Reagents and slides should be stable for 12-18 months after manufacturing when stored at temperature between +2°C and +8°C.

Anti-dsDNA IFA kit: 50 tests/kit

1. Should have Crithidia lucillae coated biochips in reaction field.
2. Should have facility to upgrade the assay with multiple substrates.
3. Should have reagent tray system for assay incubation.
4. Should be based on titreplane technique
5. Test system should be free from need of humidity chamber
6. Reagents and slides should be stable for 12-18 months after manufacturing when stored at temperature between +2°C and +8°C.

Immunoblot strip for ANA profile (ENA Profile): 16 strip/kit

1. Test system should have separate membrane strip for incubation of each serum sample.
2. Test membrane should have dsDNA, Nucleosome, Histone, Mi-2alpha, Mi-2 beta, Ku, RNP/Sm, Sm, SS-B, Ro-52, SS-A, Sp100, PML, CENP A, CENP B, gp210, PCNA, PM-Scl75, PM-Scl100, Scl-70, RP11, RP155, DFS70 coated on it.
3. Total time for performing the analysis is about 110 minutes.
4. The incubation should be performed on automated platform also.
5. All incubation steps can be processed at room temperature.
6. Each membrane should have inbuilt control band to check correct completion of incubation.

INQUIRY NO. Admin/Gen/36-02/2020-AIIMS.JDH

7. Positive and negative results should be easily and reliably differentiated from each other. The intensity of the antigen bands should correlate with the antibody titre.
8. Results should be interpretable by naked eye also.
9. The membrane strips do not contain any superfluous proteins.
10. The incubated membrane test strips must be stored for long periods.
11. The test strips should be coated with parallel lines of highly purified and biochemically characterized antigens
12. Result evaluation software should be provided free of cost.
13. Test system should have been validated/ calibrated against CDC/ WHO reference sera.
14. FDA approved test system will be preferred.

IFA based mosaic basic profile: 50 tests/kit

1. Should contain Hep20-10 cell line, rat kidney, monkey liver and rat stomach
2. Should have all above tissues/cells within same reaction field
3. Should have reagent incubation tray for test incubation
4. Should be based on the titreplane technique.
5. Test system should not have requirement for humidity chamber.
6. Test kit should contain all necessary reagents to run the test.

IFA Anti-phospholipase A2 receptor: 30 tests/kit

1. Should be able to qualitative or semiquantitative in vitro determination of human antibodies of immunoglobulin class IgG against Phospholipase A2 receptor (PLA2R) in patient samples for the diagnosis of primary membranous nephropathy.
2. Should use Biochip and Titerplane technology.
3. Presence of transfected cells and control transfected cells (EU90) as antigen substrate.
4. Reagents and slides should be stable for 12-18 months after manufacturing when stored at temperature between +2°C and +8°C.

Immunoblot strip for MPO, PR3 and GBM: 16 strips/kit

1. Test system should have separate membrane strip for incubation of each serum sample.
2. Test membrane should have MPO, PR3 and GBM coated on it.
3. Total time for performing the analysis is about 45 minutes.
4. The incubation should be performed on automated platform also.
5. All incubation steps can be processed at room temperature.
6. Each membrane should have inbuilt control band to check correct completion of incubation.
7. Positive and negative results should be easily and reliably differentiated from each other. The intensity of the antigen bands should correlate with the antibody titre.
8. Results should be interpretable by naked eye also.
9. The membrane strips do not contain any superfluous proteins.
10. The incubated membrane test strips must be stored for long periods.

INQUIRY NO. Admin/Gen/36-02/2020-AIIMS.JDH

11. The test strips should be coated with parallel lines of highly purified and biochemically characterized antigens
12. Result evaluation software should be provided free of cost.
13. Test system should have been validated/ calibrated against CDC/ WHO reference sera.
14. FDA approved test system will be preferred.

INQUIRY NO. Admin/Gen/36-02/2020-AIIMS.JDH

[On the letterhead of firm]
ANNEXURE "2"
PRICE BIDFORM

To,
Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. **“QUOTATION FOR SUPPLY OF KITS FOR IMMUNOPATHOLOGY TEST AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/36-02/2020-AIIMS.JDH” DUE ON 20.07.2020 05.00 PM** for Supply of Kits for Immunopathology Test at AIIMS Jodhpur”.

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates and as per annexure 1 as mentioned above.

S. No	Particular	Qty	Quote Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	ANA IIF	10X05						
2	ANA immunoblot	16x1						
3	ANCA IIF	10X05						
4	PLA2R IIF	10X03						
5	Mosaic basic profile	10X5						
6	IIF dsDNA	10X5						
7	MPO/PR3/GBM profile	16 strip						

Note:-

- The Bidder must quote only recommended Make.**
- The Bidder must submit the GSTIN Registration and PAN Card self-attested copy with the quotation otherwise quotation will be REJECTED.**
- The bidder must quote their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED.
- Catalog must be attached with quotation for technical evaluation.

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____