



All India Institute of Medical Sciences,
Jodhpur

Dated: 29th June, 2020

CORRIGENDUM

Invitation of quotation

for

Yasargil Permanent Aneurysm Mini-Clip

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/45-01/2020-AIIMS.JDH

Inquiry Issue Date : 25th June, 2020

Last Date of Submission : 01st July, 2020 at 05:00 PM.

The following revisions / modifications is made –

Sr No.	Item Name	Blade Length (mm)	Closing Pressure (N)	Qty.
1.	4 mm Straight	4	1.08	4 Nos.
2.	4 mm Curved	4	1.08	4 Nos.
3.	5 mm Straight	5	1.08	5 Nos.
4.	5 mm Curved	5	1.08	4 Nos.

Specification of Aneurysm Clips:-

- The instrument should be lightweight, made up of non-magnetic Titanium/Cobalt alloy with ISO 9001/CE/PIS or equivalent rating.
- The dimensions of clips may vary 5-10% from specification motioned.
- All clip should be manufactured by same manufacturer and should be compatible with the Yasargil Clip Applicator.
- The Aneurysm clip should be made of Titanium/Cobalt alloy.
- The clip should be MRI safe.
- The blade profile should have pyramid like impression.it should be atraumatic.
- The clip should ensure even pressure on the complete blade surface.
- Each clip should have individual serial no. to ensure complete tracking of manufacturing process, material, closing force, sterilization process.
- The clip should be supplied in sterile pack.
- The package of clip should mention article no., expiry date, closing force.
- The clip should be color-coded.