

### अखिल भारतीय आयुर्विज्ञान संस्थान, जोधपुर ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

Date: - 19th May 2021

### Corrigendum for Advanced Dental Chair for the Department of Dentistry

NIT Issue Date	: 23 <sup>rd</sup> April 2021
NIT No.	: Admn/Tender/02/2021-AIIMS.JDH
Pre-Bid Meeting	: 04 <sup>th</sup> May, 2021 at 03:00 PM
Earlier Last Date of Submission	: 27 <sup>th</sup> May, 2021 at 03:00 PM
Extended Last Date of Submission	: 03 <sup>rd</sup> June, 2021 at 03:15 PM
Bid opening	: 04 <sup>th</sup> June, 201 at 04:00 PM

#### The following revised and additional specification will be added:-

#### 1. Page No. 10, Motorized Programmable Dental Chair with Accessories, Point No. s: For

Should be supplied with monitor and multimedia connections, CPU with i7 processor, 8 GB RAM, 500 GB Hard Drive, wireless keyboard and mouse with Wi-Fi, LAN and Bluetooth facility and (Dell, Lenovo, HP).

#### Read

Should be supplied with monitor and multimedia connections, CPU with i7 processor, 8 GB RAM, 500 GB Hard Drive, wireless keyboard and mouse with Wi-Fi, LAN and Bluetooth facility. (Dell, Lenovo, HP or Equivalent) and should be quoted with onsite warranty.

#### 2. Page No. 11, Motorized Programmable Dental Chair with Accessories, Point No. u: For

Certification for All the Above Mentioned: Should have European CE (Four digit number)/US FDA certification.

#### Read

Certification for All the Above Mentioned: Should have US FDA/ European CE Certified and BIS certification.

#### 3. Page No. 11, Power supply, Point No. ii:

#### For

2.5 KV Servo Voltage stabilizer of appropriate rating meeting ISI Specification with each Dental Chair. (Input 160-260 V and output 220-240 V, 50 Hz). **Read** 

2.5 KV CMV (Common Mode Voltage) stabilizer of appropriate rating meeting ISI Specification with each Dental Chair. (Input 160-260 V and output 220-240 V, 50 Hz).

4. Page No. 11, Point No. B, Head Line:-For
B. COMPRESSOR UNIT AND SUCTION UNIT (COMBINED)
Read
COMPRESSOR UNIT AND SUCTION UNIT (COMBINED): Quantity/Unit (02 in Number)

# 5. Page No. 11, B. COMPRESSOR UNIT AND SUCTION UNIT (COMBINED), Point No. 1: For

Modular Medical grade Durr Compressor and Suction combined in a Single Housing. Read

Modular Medical grade (Durr or Equivalent) Compressor and Suction combined in a Single Housing.

### 6. Page No. 11, B. COMPRESSOR UNIT AND SUCTION UNIT (COMBINED), Point NO. 2: For

Combined Noiseless Single Unit should be able to operate three (03) Dental Chairs simultaneously.

#### Read

Combined Noiseless Single Unit should be able to operate four (04) Dental Chairs simultaneously.

7. Page No. 12, B. COMPRESSOR UNIT AND SUCTION UNIT (COMBINED), Point No. 5: For

Absolutely 100% Oil free, dry, hygienic compressed air with membrane drying System. **Read** 

Absolutely 100% Oil free, dry, hygienic compressed air with membrane drying System/ desiccant dryer/ membrane dryer.

# 8. Page No. 12, B. COMPRESSOR UNIT AND SUCTION UNIT (COMBINED), Point No. 17: For

The equipment must bear European CE (Four digit number)/ US FDA conformity marking. **Read** 

The equipment must bear US FDA/ European CE Certified and BIS conformity marking.