



Date: - 17th June, 2020

Corrigendum
for
Corneal topography for the Department of
Ophthalmology

NIT Issue Date	: 20 th May, 2020
NIT No.	: Admn/Tender/05/2020-AIIMS.JDH
Pre-Bid Meeting	: 01 st June, 2020 at 03:45 PM
Last Date of Submission	: 29 th June, 2020 at 03:00 PM
Bid opening	: 30 th June, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 06, Point No. 01:

For

Measuring Technology -Placido disk + Scheimpflug camera

Read as:

Measuring Technology should be based on Placido disk/ Scheimpflug camera.

2. Page No. 06, Point no. 5:

For

Analysed Points over 100,000, Measured points over 6.200

Read as:

Analysed Points over 100,000 or Measured points over 6.200