



Date: - 21<sup>st</sup> June, 2018

Corrigendum  
For  
Tender for  
Dermatomes (Power Drive) for the Department of  
Burns and Plastic Surgery

NIT Issue Date	: 09 <sup>th</sup> April, 2018
NIT No.	: Admn/Tender/06/2018-AIIMS.JDH
Pre-Bid Meeting	: 19 <sup>th</sup> April, 2018 at 03:15 PM
Earlier Last Date of Submission	: 25 <sup>th</sup> June, 2018 at 03:00 PM
Extended Last Date of Submission	: 04 <sup>th</sup> July, 2018 at 03:00 PM
Bid opening	: 05 <sup>th</sup> July, 2018 at 03:15 P.M

**The following revised and additional specification will be added:-**

**1. Page No. 11, 1<sup>st</sup> Point:**

**For**

The thickness control adjustment ranges: 0 to 0.030 in. (0.75 mm) and increment is 0.002 in. (0.050 mm).

**Read**

The thickness control adjustment ranges: 0.004 in. (0.10 mm) to 0.048 in. (1.20mm) and increment is 0.004 in. (0.10mm)

**2. Page No. 11, 2<sup>nd</sup> Point:**

**For**

Individual graft widths: 1 in., 2 in., 3 in. and 4 in. (2.5 cm, 5.1 cm, 7.6 cm, and 10.2 cm) are obtained with four width plates. Two stainless steel machine screws secure the plates to the underside of the instrument. The plates are easily fastened and removed with the screwdriver provided.

**Read**

0.4in., 0.8 in., 1.2 in., 1.6 in., 2.0 in., 2.4 in., 2.8 in., and 3.2 in. (1 cm, 2 cm, 3 cm, 4 cm, 5 cm, 6 cm, 7 cm, and 8 cm) are obtained with eight width plates.

Two stainless steel machine screws secure the plates to the underside of the instrument. The plates are easily fastened and removed with the screwdriver provided.

**3. Page No. 11, Addition to the 4<sup>th</sup> point:**

3000 MAH Li-ion Rechargeable battery

**4. Page No. 11, 5<sup>th</sup> point:**

**For**

Weight: 2.1 lb. (.95 kg)

Length: 5.2 in. (13.2 cm)

Width: 8.5 in. (21.6 cm)

**Read**

Weight: range of 0.80 kg- 1.30 kg

Length: 25 cm to 30 cm

Width: 10 cm to 12 cm

**5. Page No. 11 , Addition of a point:**

Each Dermatome should be supplied with 250 nos. of blades.