



Date: - 18th December, 2018

Corrigendum
For
Ultra - Centrifuge for the Department of Microbiology

NIT Issue Date	: 22 nd October, 2018
NIT No.	: Admn/Tender/101/2018-AIIMS.JDH
Pre-Bid Meeting	: 30 th October, 2018 at 05:00 PM
Earlier Last Date of Submission	: 13 th December, 2018 at 03:00 PM
Extended Last Date of Submission	: 27 th December, 2018 at 03:00 PM
Bid opening	: 28 th December, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, Point No. 1, Under heading “Technical”:**
For
Bench top, compact, Refrigerated
Read
Bench top/Floor model, compact, Refrigerated
- 2. Page No. 10, Point No. 2, Under heading “Technical”:**
For
Temperature setting: 0° to 4° C.
Read
Temperature setting: 0° to 40° C.
- 3. Page No. 10, Point No. 5, Under heading “Technical”:**
For
Speed upto 15000 RPM
Read
Speed-150000 RPM