



Date: - 14th February, 2017

Corrigendum
for
Ambulatory Blood Pressure Machine for the
Department of Cardiology

NIT Issue Date	:	09 th January, 2017
NIT No.	:	Admn/Tender/102/2016-AIIMS.JDH
Pre Bid Meeting held on	:	18 th January, 2017 at 03:00 PM
Earlier Last Date of Submission	:	13 th February, 2017 at 03:00 PM
Extended Last Date of Submission	:	28 th February, 2017 at 03:00 PM
Bid Opening	:	01 st March, 2017 at 04:00 PM

The following revised and additional specification will be added:-

1. Page No. 11, In technical specification, Para No. 07:

For

Memory Capacity 300 measurement maximum

Read

Memory Capacity ≥ 300 measurements

2. Page No. 11, In technical specification, Para No. 08:

For

Power Source Three AA Alkaline or NiCd (Batteries not included)

Read

Power Source Two or Three AA Alkaline or NiCd (Batteries included)

3. Page No. 11, In technical specification, Para No. 12:

For

Data Output RS-232C (direct)

Read

Data Output RJ 45 or equivalent ports

4. Page No. 11, In technical specification, Para No. 13:

For

Cuffs	Adult Cuffs, Left Arm	(7.9" - 12.2" (20-31cm)) included
	Large Cuffs, Left Arm	(11" - 14.2" (28-36cm)) Optional
	Small Cuffs, Left Arm	(5.9" - 8.7" (15-22cm)) Optional
	Adult Cuffs, Right Arm	(7.9" - 12.2" (20-31cm)) optional
	Large Cuffs, Right Arm	(11" - 14.2" (28-36cm)) Optional

Read

	Adult Cuffs, Left Arm	(7.9" - 12.2" (20-31cm) (Included)
	Large Cuffs, Left Arm	(11" - 14.2" (28-36cm) (Included)
Cuffs	Small Cuffs, Left Arm	(5.9" - 8.7" (15-22cm) (Included)
	Adult Cuffs, Right Arm	(7.9" - 12.2" (20-31cm) (Included)
	Large Cuffs, Right Arm	(11" - 14.2" (28-36cm) (Included)

5. Page No. 11, In technical specification, Para No. 14:

For

Doctor Pro software

Read

Doctor Pro software Appropriate Software should be provided

6. Page No. 11, In technical specification, After Para No. 23:

Added Para No. 24:

Hardware Requirements Analyzer software should be provided.

**Administrative Officer
AIIMS, Jodhpur**