



Date: - 14th February, 2017

Corrigendum
for
Defibrillator (High End) for the Department of
Cardiology

NIT Issue Date	:	09 th January, 2017
NIT No.	:	Admn/Tender/103/2016-AIIMS.JDH
Pre Bid Meeting held on	:	18 th January, 2017 at 03:00 PM
Earlier Last Date of Submission	:	13 th February, 2017 at 03:00 PM
Extended Last Date of Submission	:	28 th February, 2017 at 03:00 PM
Bid Opening	:	01 st March, 2017 at 04:00 PM

The following revised and additional specification will be added:-

1. Page No. 11, In technical specification, Point No. 03:

For

Should have factory integrated compensation for chest impedance for a range of 25 to 150 ohms.

Read

Should have factory integrated compensation for chest impedance for a range of 15 to 300 ohms.

2. Page No. 11, In technical specification, Point No. 05:

For

Should have charging time of less than 5 seconds for maximum energy. Charging indicator should be present.

Read

Should have charging time of less than 8 seconds for maximum energy. Charging indicator should be present.

3. Page No. 11, In technical specification, Point No. 06:

For

Should have bright TFT colour display 8" or more for viewing messages and ECG waveform of 4 seconds.

Read

Should have bright TFT color display 6" or more for viewing messages and ECG waveform of 4 seconds.

4. Page No. 11, In technical specification, Point No. 07:

For

Should have external paddles with paddle contact indicators. Single adult and paediatric paddles should be available.

Read

Should have external paddles. Single adult and pediatric paddle should be available.

5. Page No. 11, In technical specification, Point No. 09:

For

Should have a battery capable of usage for at least 5 hours or 50 discharges

Read

Should have a battery capable of usage for at least 4 hours or 100 discharges

6. Page No. 11, In technical specification, System Configuration Accessories, spares and consumables Point No. 03:

For

Deleted

Read

Removed

**Administrative Officer
AIIMS, Jodhpur**