Date: - 11th January, 2019

Corrigendum

For

CCU Beds +Cardia Table +Side Lockers for the Department of Cardiology

NIT Issue Date : 01st December, 2018

NIT No. : Admn/Tender/12-2/2018-AIIMS.JDH

Pre-Bid Meeting : 10th December, 2018 at 03:45 PM

Earlier Last Date of Submission : 14th January, 2019 at 03:00 PM

Extended Last Date of Submission: 30th January, 2019 at 03:00 PM

Bid opening : 31st January, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Annexure I, S No. 1, Item: CCU Beds +Cardia Table +Side Lockers, in technical specification, point No. 2

For

Dimensions of bed

• Length x Width: 2280 mm x 950 mm or better

• Mattress Size: 840 x 2020mm

• Permissible load capacity: 250 kg or more

Read

Dimensions of bed

• Length x Width: 2280 mm x 950 mm or better

• Mattress Size: 840 x 2060 mm

• Permissible load capacity: 250 kg or more

2. Page No. 10, Annexure I, S No. 1, Item: CCU Beds +Cardia Table +Side Lockers, in technical specification, point No. 3(a)

For

Height: min. 485 mm or less (electrically) max. 820 mm or better (electrically)

Read

Height: min. 485 mm or less (electrically) max. 880 mm or better (electrically)

3. Page No. 10, Annexure I, S No. 1, Item: CCU Beds +Cardia Table +Side Lockers, in technical specification, point No. 3(d)

For

Shank Rest Adjustment: 16 degrees or better (pneumatically/Motorised)

Read

Shank Rest Adjustment: 22 degrees or better (pneumatically)

4. Page No. 10, Annexure I, S No. 1, Item: CCU Beds +Cardia Table +Side Lockers, in technical specification, point No. 3(e)

For

Trendelenburg: 12 degrees or better (electrically)

Read

Trendelenburg: 15 degrees or better (electrically)

5. Page No. 10, Annexure I, S No. 1, Item: CCU Beds +Cardia Table +Side Lockers, in technical specification, point No. 3(f)

For

Anti-Trendelenburg: 12 degrees or better (Electrically)

Read

Anti-Trendelenburg: 13 degrees or better (Electrically)

6. Page No. 10, Annexure I, S No. 1, Item: CCU Beds +Cardia Table +Side Lockers, in technical specification, point No. 10

For

It should have facility of bed length extension for large patients up to 18 cm or more. Bed should have inbuilt retractable linen shelf

Read

It should have facility of bed length extension for large patients up to 20 cm. Bed should have inbuilt retractable linen shelf