



Date: - 04th April, 2018

Corrigendum
For
Tender for
Image Guided Navigation System for the Department
of ENT

| | |
|----------------------------------|--|
| NIT Issue Date | : 09 th February 2018 |
| NIT No. | : Admn/Tender/165-2/2017-AIIMS.JDH |
| Pre-Bid Meeting | : 20 th February 2018 at 03:30 PM |
| Earlier Last Date of Submission | : 10 th April 2018 at 03:00 PM |
| Extended Last Date of Submission | : 17 th April 2018 at 03:00 PM |
| Bid opening | : 18 th April 2018 at 03:15 P.M |

The following revised and additional specification will be added:-

1. Page No. 10, Annexure –I, In technical Specification, Para No. 04:

For

The system should have facility of keeping optical camera and viewing system together or separately to allow optimal use of O.T space, the system should have two monitors or two cart system.

Read

The system should have facility of keeping optical camera and viewing system together or separately to allow optimal use of O.T space, the system should have two monitors and two cart system.

2. Page No. 10, Annexure –I, In technical Specification, Para No. 05:

For

The surgeon monitor should be high resolution (1920x1200, 60Hz) with a viewable size of 20” or more widescreen.

Read

The surgeon monitor should be high resolution (1920x1200, 60Hz) with a viewable size of ≥ 23 ” inch and the display should be visible at angles at least up to 80^o or more.

3. Page No. 11, Annexure –I, In technical Specification, Para No. 17:

For

Cart/Movable trolley for housing should have a storage capability.

Read

Cart/Movable trolley for housing should have an inbuilt storage capability.



4. Page No. 11, Annexure –I, In technical Specification, Para No. 22:

For

European CE and US FDA certification

Read

Company should provide European CE or US FDA approval certificate for the quoted model.

5. Page No. 12, Annexure –I, In technical Specification, Point No. 19:

For

Should have facility to integrate third party instruments - required adapter size should be for small, medium, large, extra-large, and conical and for cylindrical instruments. Also a set of disposable instruments set for at least 200 surgeries should be provided.

Read

Should have facility to integrate third party instruments, with the universal adapter for various sizes and shapes of instruments. If disposable, minimum 10 sets of disposable instruments/consumables should be provided with the unit.

6. Page No. 12, Annexure –I, In technical Specification, In Last Point:

For

The company will provide software/ technological upgrades to the entire supplied unit for next five years free of cost.

Read

The company will provide software/ technological updates to the entire supplied unit for next five years free of cost.

7. Page No. 12, Annexure –I, In technical Specification, After Point No. 21:

Added Point No. 22:

The system should have the options of both flat Emitter and side Emitter to support the workflow of the surgeries and both should be provided. The flat panel electromagnetic generator is one which can be positioned under patient's head and over the table in order to minimize table metal interference and side mounted electromagnetic generator is one which can be attached to the table side rail.