



Date: -29th November, 2017

Corrigendum
For
Tender for
Flexible Videobronchoscope (White light) for the
Department of Pulmonary Medicine

NIT Issue Date	: 15 th September, 2017
NIT No.	: Admn/Tender/179/2017-AIIMS.JDH
Pre-Bid Meeting	: 27 th September, 2017 at 04:15 PM
Earlier Last Date of Submission	: 30 th November, 2017 at 03:00 PM
Extended Last Date of Submission	: 19 th December, 2017 at 03:00 PM
Bid opening	: 20 th December, 2017 at 03:30 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 08:

For

Depth of Field should be 3-100mm or better.

Read

Depth of Field should be 3-50mm or better.