Date: - 31st January, 2018

Corrigendum

For

Tender for

CO₂ Incubator for the Department of Microbiology

NIT Issue Date : 09th October, 2017

NIT No. : Admn/Tender/206/2017-AIIMS.JDH

Pre-Bid Meeting : 24th October, 2017 at 03:30 PM

Earlier Last Date of Submission : 05th February, 2018 at 03:00 PM

Extended Last Date of Submission : 20th February, 2018 at 03:00 PM

Bid opening : 21st February, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, point no. 1

For

Air jacket with internal capacity: 150 L (Approx)

Read

Air jacket with internal capacity: 160 L or more

2. Page No. 10, point no. 10

For

CO2 Range- 0-20%; CO2 Accuracy: +/- 0.5%; CO2 Inlet pressure 1.5 bars (app) and fast recovery after opening door.

Read

CO2 Range- 0-20%; CO2 Accuracy: +/- 0.5%; CO2 Inlet pressure 0.8 to 1.5 bars (app) and fast recovery after opening door.

3. Page No. 10, point no. 13

For

72-Hour Data Storage for CO2 concentration, temperature, alarms and door openings should be automatically recorded for on-screen displays.

Read

72-Hour Data Storage for CO2 concentration, temperature, alarms and door openings should be automatically recorded for on-screen displays. With data lock facility.

4. Page No. 10, point no. 14

For

Data output through RS Port.

Read

Data output through USB.

5. Page No. 10, point No. 16

For

CO2 cylinders 2 nos. (capacity at least 30 kg) with regular (at least one) compatible to machine part

Read

CO2 cylinders 2 nos. (capacity at least 30 kg) with regular (at least one) compatible to machine part with CO_2 automatic cylinder change over.

6. Page No. 10, point no. 29

For

IR sensor for CO2 measurement

Read

TC/IR sensor for CO2 measurement

7. Page No. 10, point no. 31

For

Decontamination at 180° C should be provided.

Read

Decontamination at 180°C or more should be provided.