Date: - 03rd August 2016

Corrigendum

for

Orthopaedic Bed with Traction Pulley

NIT No. : Admn/Tender/21/2016-AIIMS.JDH

NIT Issue Date : 23rd June, 2016

Pre Bid Meeting : 04th July, 2016 at 03:00 PM

Last Date of Submission : 20th July, 2016 at 03:00 PM

Extended Last Date of Submission : 08th August, 2016 at 03:00 PM

Revised Last Date of Submission : 22nd August, 20165 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Page No. 10, Chapter-II Specification and Allied Technical Specification, Parameters and Technical Specifications for Executing the Work:, Point No. 01:

For

All dimensions shall be read in metric unit only. Tolerance in overall dimensions upto \pm 5%. No other deviations from specifications shall be allowed.

Read

All dimensions shall be read in metric unit only. Tolerance in overall dimensions upto -2% to +5%. No other deviations from specifications shall be allowed.

2. Page No. 10, Chapter-II Specification and Allied Technical Specification, Parameters and Technical Specifications for Executing the Work:, General Eligible Criteria, Point No. 02:

For

Manufacturer should have valid European CE, ISO 9001:2008, ISO 14001:2004, and OHSAS 18001:2007 certificate. Dealers participating should enclose certificate from their parent manufacturer company.

Read

Manufacturer should have valid CE, ISO 9001:2008, ISO 14001:2004, and OHSAS 18001:2007 certificate. Dealers participating should enclose certificate from their parent manufacturer company.

3. Page No. 14, Chapter-V, TECHNICAL SPECIFICATION FOR FOLWER BED WITH RAILS, Bullet No. 08:

For

Should have easily removable head and foot panels made up of SS with four corner buffers.

Should have easily removable head and foot panels made up of moulded polymer panels mounted in SS frame with four corner buffers.

4. Page No. 14, Chapter-V, TECHNICAL SPECIFICATION FOR FOLWER BED WITH RAILS, Bullet No. 13:

For

Each bed must be quoted with one no, four section mattress of dimension (2000 mm X 900 mm) with washable cover of lasting quality. The mattress should be made of high density PU foam of 100 mm thickness.

Read

Each bed must be quoted with one no, four section mattress of dimension (2000 mm X 900 mm) with washable cover of lasting quality. The mattress should be made of 40 density and PU foam of 100 mm thickness. The Mattress cover should be water resistant and fire retardant.

Administrative Officer AIIMS, Jodhpur