



Date: - 13<sup>th</sup> October, 2018

Corrigendum  
For  
Tender for  
Dialysis Reprocessing unit for the Department of  
Nephrology

NIT Issue Date	: 10 <sup>th</sup> August, 2018
NIT No.	: Admn/Tender/225-3/2017-AIIMS.JDH
Pre-Bid Meeting	: 20 <sup>th</sup> August, 2018 at 03:45 PM
Earlier Last Date of Submission	: 15 <sup>th</sup> October, 2018 at 03:00 PM
Extended Last Date of Submission	: 26 <sup>th</sup> October, 2018 at 03:00 PM
Bid opening	: 27 <sup>th</sup> October, 2018 at 03:15 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, In Technical Specification, Point No. 11:**  
**For**  
Sterilant volume indicator cartridge  
**Read**  
Sterilant volume indicator
- 2. Page No. 11, In Technical Specification, Point No. 13:**  
**For**  
Facility to check fiber bundle leakage at 250mm hg  
**Read**  
Facility to check fiber bundle leakage at -250mm hg