



Date: - 26<sup>th</sup> April, 2018

Corrigendum  
For  
Tender for  
Defibrillator Machine for OPD and IPD services

NIT Issue Date	: 16 <sup>th</sup> March, 2018
NIT No.	: Admn/Tender/257/2017-AIIMS.JDH
Pre-Bid Meeting	: 21 <sup>st</sup> March, 2018 at 03:00 PM
Earlier Last Date of Submission	: 07 <sup>th</sup> May, 2018 at 03:00 PM
Extended Last Date of Submission	: 10 <sup>th</sup> May, 2018 at 03:00 PM
Bid opening	: 11 <sup>th</sup> May, 2018 at 03:15 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, Specification, Point No. 05:**  
**For**  
It should have facility of manual defibrillation and upgradable to External Pacing.  
**Read**  
It should have facility of manual defibrillation. Out of all (38 No.) the Machines, 12 machines needs to be supplied with facility of external Pacing and remaining machines should be upgradable to External Pacing. Rate of such upgarde to External pacing needs to be quated saperatly.
- 2. Page No. 11, Specification, Point No. 07:**  
**For**  
It should have in built 3 channel ECG recorder.  
**Read**  
It should have in built 2 or more channel ECG recorder.
- 3. Page No. 11, Specification, Point No. 10:**  
**For**  
Should have synchronous cardioversion.  
**Read**  
Facility of ECG synchronization, Standard patient cable ECG; minimum leads I, II and III, Paddle ECG
- 4. Page No. 11, Specification, Point No. 12:**  
**For**  
It should have biphasic mode, with range of 1 to 200 joules, charge time less than 5 seconds.  
**Read**  
It should have biphasic mode, with range of 2 to 200 joules, charge time less than 5 seconds.

- 5. Page No. 11, Specification, Point No. 14:**  
**For**  
Should have battery capacity indicator on battery.  
Read  
Should have battery capacity indicator on Monitor
- 6. Page No. 11, Specification, Point No. 15:**  
**For**  
It should meet international safety standards, US-FDA and CE certification  
Read  
It should meet international safety standards, US-FDA/ European 4 digit CE certification
- 7. Page No. 11, Specification, Point No. 16:**  
**For**  
It should be capable of transferring data to PC through USB flash memory.  
Read  
It should be capable of transferring data to PC through USB flash memory/ SD Card/  
USB cable.
- 8. Page No. 11, Specification, Point No. 17, Sub Toggle No. 03:**  
**For**  
Patient cables-01  
Read  
Two sets of ECG cables
- 9. Page No. 11, Specification, Point No. 19:**  
**For**  
It should be capable of operating on mains power Input of 120-240V, 50/60Hz.  
Read  
It should be capable of operating on mains power Input of 220-240VAC, 50Hz fitted with  
Indian plug
- 10. Page No. 11, Specification, After Point No. 19:**  
**Added Point No. 20:**  
Should give rate of all spares and accessories including consumables, which will be fixed  
for 2 years atleast.
- 11. Page No. 11, Specification, After Point No. 20:**  
**Added Point No. 21:**  
Technical Bid Compliance statement: The vendor must provide, a comparative chart of  
the required technical specification and technical specification of the quoted product in  
tabular format . Vendor must specify marketing name of quoted product. The vendor must  
give the relevant page number and paragraph number of their literature/Equipment Data  
Sheet regarding conforming technical information in the technical bid table. Merely stating  
“complies” or “meets requirement” will lead to assumption that the quoted product does  
not have the required feature.