**Date: - 26<sup>th</sup> April, 2018** 

### Corrigendum

### For

### Tender for

### ECG Machines with Accessories for OPD and IPD

NIT Issue Date : 16<sup>th</sup> March, 2018

NIT No. : Admn/Tender/260/2017-AIIMS.JDH

Pre-Bid Meeting : 21<sup>st</sup> March, 2018 at 03:00 PM

Earlier Last Date of Submission : 07<sup>th</sup> May, 2018 at 03:00 PM

Extended Last Date of Submission: 10th May, 2018 at 03:00 PM

Bid opening : 11<sup>th</sup> May, 2018 at 03:15 P.M

### The following revised and additional specification will be added:-

#### 1. Page No. 11, Point No. 03, Technical Specification, Sub Point No. 3.4:

For

Should have a storage memory of at least 100 ECGs with easy transfer by modem and data card.

### Read

Should be capable to transfer stored data to USB/Data card/ Pendrive

### 2. Page No. 11, Point No. 03, Technical Specification, Sub Point No. 3.10:

For

Should have battery capacity of at least 30 ECGs or 30 minutes of continuous rhythm recording on single charge.

#### Read

Should have rechargeble inbuilt battery with capacity of at least 100 or more ECGs or 90 minutes or more of continuous rhythm recording on single charge.

### 3. Page No. 11, Point No. 03, Technical Specification, Sub Point No. 3.12:

For

Should display ECG on LCD/TFT Display of 640x480 pixel resolution.

#### Read

Should display 12 chennel ECG on LCD/TFT Display of 640x480 pixel or higher resolution.

### 4. Page No. 11, Point No. 03, Technical Specification, Sub Point No. 3.14:

### Added Point NO. 3.15:

Should have sampling rate of >15000 sample/sec/chennel

## 5. Page No. 11, Point No. 04, System Configuration Accessories, spares and consumables, After Sub Point No. 4.6:

Added Point NO. 4.7:

Preferablly OEM or equalant Branded Trolly as specified below- One

# 6. Page No. 11, Point No. 04, System Configuration Accessories, spares and consumables, After Sub Point No. 4.7:

Added Point NO. 4.8:

Trolley should made of Stainless Steel / Powder coated frame with SS 304 grade Top

# 7. Page No. 11, Point No. 04, System Configuration Accessories, spares and consumables, After Sub Point No. 4.8:

Added Point NO. 4.9:

Should be a 3-shelf (including the top) cart, one with a drawer for storing the accessories and consumables.

# 8. Page No. 11, Point No. 04, System Configuration Accessories, spares and consumables, After Sub Point No. 4.9:

Added Point NO. 4.10:

Should have four superior castors (two with brakes)

## 9. Page No. 11, Point No. 04, System Configuration Accessories, spares and consumables, After Sub Point No. 4.10:

Added Point NO. 4.11:

Trolley should have at least 30" height and the shelves should have sufficient space for storing the accessories.

# 10. Page No. 11, Point No. 04, System Configuration Accessories, spares and consumables, After Sub Point No. 4.11:

Added Point NO. 4.12:

Top shelves shall be surrounded by railing.

# 11. Page No. 11, Point No. 04, System Configuration Accessories, spares and consumables, After Sub Point No. 4.12:

Added Point NO. 4.13:

Trolley should have a suitable cable arm firmly affixed having holderfor ECG cables while not in use

#### 12. Page No. 12, Point No. 07, Standards, Safety and Training, Sub Point No. 7.1:

For

Should be US FDA and European CE, approved product.

Read

Should be US FDA/4 digit European CE, approved product.

## 13. Page No. 12, Point No. 08, Documentation, After Sub Point No. 8.6: Added Point NO. 8.7:

Technical Bid Compliance statement: The vendor must provide, a comparative chart of the required technical specification and technical specification of the quoted product in tabular format. Vendor must specify marketing name of quoted product. The vendor must give the relevant page number and paragraph number of their literature/Equipment Data Sheet regarding conforming technical information in the technical bid table. Merely stating "complies" or "meets requirement" will lead to assumption that the quoted product does not have the required feature.