

## अखिल भारतीय आयुर्विज्ञान संस्थान, जोधपुर ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

Date: - 29<sup>th</sup> June, 2018

# Corrigendum For

Tender for

# Laparoscopic Instruments – Thoracoscopic for the Department of Surgical Gastroenterology

NIT Issue Date	:	09 <sup>th</sup> April, 2018
NIT No.	:	Admn/Tender/28/2018-AIIMS.JDH
Pre-Bid Meeting	:	25 <sup>th</sup> April, 2018 at 03:00 PM
Earlier Last Date of Submission	:	25 <sup>th</sup> June, 2018 at 03:00 PM
Extended Last Date of Submission	:	09 <sup>th</sup> July, 2018 at 03:00 PM
Bid opening	:	10 <sup>th</sup> July, 2018 at 03:15 P.M

## The following revised and additional specification will be added:-

1. Page No. 11, 1<sup>st</sup> Sub point of Laparoscopic needle holder:

For

Disposable Laparoscopic Stitching Device With Capability To Transfer Needle From One Limb To The Other Limb By Transfer Mechanism Hand Operated With 10mm Shaft **Read** 

Reusable Laparoscopic Stitching Device With Capability To Transfer Needle From One Limb To The Other Limb By Transfer Mechanism Hand Operated With 10mm Shaft

- 2. Page No. 11, 1<sup>st</sup> sub point of clip applicator: For Reusable Endo Clip Applier With Rotating Shaft For Medium Clips (Lt 200) Read Reusable Endo Clip Applier With Rotating Shaft For Medium Clips.
- Page No. 11, 2<sup>nd</sup> sub point of clip applicator: For Reusable Endo Clip Applier With Rotating Shaft For Medium Large Clips (Lt 300). Read
  Reusable Endo Clip Applier With Rotating Shaft Medium Large Clips
  - Reusable Endo Clip Applier With Rotating Shaft Medium Large Clips .
- Page No. 11, 3<sup>rd</sup> sub point of clip applicator For Reusable Endo Clip Applier With Rotating Shaft For Large Clips (Lt 400)

Read

Reusable Endo Clip Applier With Rotating Shaft For Extra Large Clips.

### 5. Page No.11, 4<sup>th</sup> sub point of clip applicator

For

Reusable Hemostatic Clip Applier For Laparoscopic Surgery For Application Of Hemostatic Locking Non Absorbable Clips Of Sizes Medium Large, Large And Extra Large, 5 Mm Diameter 32 Cm Curved

Read

Reusable Hemostatic Clip Applier For Laparoscopic Surgery For Application Of Hemostatic Locking Non Absorbable Clips Of Sizes Medium Large, Large And Extra Large, 5 Mm Diameter 32 cm

## 6. Page No. 11, 6<sup>th</sup> sub point of clip applicator

#### For

Bipolar High Frequency Cord Compatible With Bipolar Electro Surgical Generator And Bipolar Laparoscopic Hand Instruments

#### Read

No electrosurgical generator required only compatible Bipolar cord and bipolar laparoscopic hand instruments

## 7. Page No. 12, 9<sup>th</sup> sub point of Trocars

#### For

Disposable Opaque Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5 Mm, 75/100 Mm Length

#### Read

Disposable Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5 Mm, 75/100 Mm Length

## 8. Page No. 12, 11<sup>th</sup> sub point of Trocars

#### For

Disposable Opaque Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5-12 Mm, 75/100 Mm Length

Read

Disposable Body Bladeless Trocar With Converter Less Insertion Of Instruments With Different Diameters, Having Insuffulation Port And Ribbed Cannula, With Obturator Having Safety Shield. 5-12 Mm, 75/100 Mm Length

#### 9. Page No. 12, Addition of paragraphs

Haemostatic locking non absorbable clips medium large, large, extra large

#### 10. Page No. 12, Addition of paragraphs

Laparoscopic Reusable haemostatic lock clip remover. Quantity 1

#### 11. Page No. 12, Addition of paragraphs

The all instruments mentioned should be quoted and the 80% must be from the same manufacturer and 20% may from  $3^{rd}$  party but the responsibility of the warranty will be of the bidder and all instruments should be as per quality standard asked in the tender.

#### 12. Page No. 12, Addition of paragraphs

Instruments should be European CE (with four digit notified body no.) or US FDA for the quoted instruments