Date: - 17th September, 2018

Corrigendum For

ICU Beds for Morbid Obese Patient with Weighing Scale for the Department of Anesthesiology

NIT Issue Date : 14th August, 2018

NIT No. : Admn/Tender/46-2/2018-AIIMS.JDH

Pre-Bid Meeting : 24th August, 2018 at 03:15 PM

Earlier Last Date of Submission : 24th September, 2018 at 03:00 PM

Extended Last Date of Submission: 03rd October, 2018 at 03:00 PM

Bid opening : 04th October, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, Under heading "ICU Beds for Morbid Obese Patient with Weighing Scale":

Point	For	Read
No.		
2	Straight/Split Side rails with fold away movement with illuminated hand held control panel for patient & care giver.	Straight/Split Side rails with fold away movement with illuminated hand held/Embedded control panel for patient & care giver.
4	Straight metallic collapsible side rails.	Deleted
11	Should have auto lock provision of controls in case bed is not getting used for some time	Deleted
13	Non powered instant CPR (Quick release) & Electric CPR (mattress platform become flat & goes to minimum height for better patient access)	Non powered instant CPR (Quick release) & Electric CPR (mattress platform become flat)
15	Hand held attendant/nurse control panel at foot end.	Hand held /Embedded attendant/nurse control panel at foot end.
19	10-22-32 cm step wise bed extension	Step wise bed extension facility.

2. Addition of a point 25th after 24th under heading: ICU Beds for Morbid Obese Patient with Weighing Scale:

Integrated weighing scale with accuracy +/- 500 gram

3. Addition of a point 26th after 25th under heading: ICU Beds for Morbid Obese Patient with Weighing Scale:

Must qualify IEC60601-2-52:2009 by approved European or US notifying body for international safety standard for electric hospital bed, Shock protection Class-1

4. Addition of a point 27th after 26th under heading: ICU Beds for Morbid Obese Patient with Weighing Scale:

Digital/Analog angle indicator for back rest/knee break

5. Page No. 11, Point No. 4 under heading "Accessories":

For

To be supplied with pressure reduction /shear reduction Foam mattress of height 27cm, of 40 density, covered water proof, antibacterial and flame retardant cover. Mattress should be Medical graded and OEM with Certification.

Read

To be supplied with pressure reduction /shear reduction Foam mattress of minimum height 15cm, of 40 density, covered water proof, antibacterial and flame retardant cover. Mattress should be Medical graded and OEM with Certification.

6. Page No. 11, Point No. 8 under heading "Accessories":

For

X-ray and C-Arm permeable back-rest with X-ray cassette holder.

Read

X-ray permeable back-rest with X-ray cassette holder.