



Date: -23<sup>rd</sup> June, 2017

Corrigendum  
For  
Tender for  
Non Invasive Cerebral Oximeter Monitor for the  
Department of Anaesthesiology

NIT Issue Date	: 11 <sup>th</sup> April, 2017
NIT No.	: Admn/Tender/52/2017-AIIMS.JDH
Pre-Bid Meeting	: 24 <sup>th</sup> April, 2017 at 02:15 PM
Earlier Last Date of Submission	: 27 <sup>th</sup> June, 2017 at 03:00 PM
Extended Last Date of Submission	: 12 <sup>th</sup> July, 2017 at 03:00 PM
Bid opening	: 13 <sup>th</sup> July, 2017 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, In technical specification, Point No. 4, In Operationally it must have following, Sub Point No. i:**  
**For**  
Range of rSO<sub>2</sub> 14 – 96(updated every 4 – 8 seconds)  
**Read**  
Range of rSO<sub>2</sub> 14 – 96(updated within 8 seconds)
- 2. Page No. 11, In technical specification, Point No. 4, In Operationally it must have following, Sub Point No. iv:**  
**For**  
Diagnostics Automatic self-test Safety Class Continuous Operation  
**Read**  
Deleted