



Date: - 25th January, 2020

Corrigendum-2
for
HD Video Colposcope and Image management system
for the Department of Obstetrics and Gynaecology

NIT Issue Date	: 26 th November, 2020
NIT No.	: Admn/Tender/59/2020-AIIMS.JDH
Pre-Bid Meeting	: 07 th December, 2020 at 03:15 PM
Earlier Last Date of Submission	: 28 th January, 2020 at 03:00 PM
Extended Last Date of Submission	: 09 th February, 2021 at 03:00 PM
Bid opening	: 10 th February, 2021 at 03:15 P.M

The following revised and additional specification will be added: -

1. Page no. 10, point no. 3

For

The video Colposcope must have magnification from min. 1x to max. 45X

Read as-

The video Colposcope must have magnification from **min.1X to max. 55X**

2. Page no. 10, point no. 6

For

Auto focus range should be up to 30-40cm

Read as-

Auto focus range should be up to 30-40cm and **intensity of light can be increased or decreased**

3. Page no. 10, point no. 9

For

There must be Electronic Green Filter in the hand-held unit without decrease in Illumination.

Read as-

There must be Electronic Green Filter and **Blue filter** in the hand-held unit without decrease in Illumination

4. Page no. 10, last point-

For

Company has to provide Desktop / laptop Computer with: CPU Intel Core i10, RAM 32 GB, Hard Disk 500GB, Window Home Basic or better, colored Printer, DVD Writer

Sturdy Fiber body trolley- where computer and printer can be fitted

Read as-

Company has to provide Desktop Computer with: Processor - Intel Core i7-7700(3.6 Ghz , 8MB cache, 4 cores), integrated graphics- Intel HD graphics 530 or higher, RAM 16GB, RAM expandability 32 GB or higher, Hard Drive 1TB, Monitor 19 inches or higher, Processor generation 7th, VGA & HDMI port available, Standard keyboard and mouse, and colored Printer,

Sturdy Fiber body trolley with wheels-where Desktop computer and printer can be fitted