



Date: - 27<sup>th</sup> January 2020

Corrigendum  
for  
Surgical Micro Drill & Saw System for the  
Department of Dentistry

|                                  |  |
|----------------------------------|--|
| NIT Issue Date                   | : 04 <sup>th</sup> November, 2019              |
| NIT No.                          | : Admn/Tender/88-2/2018-AIIMS.JDH              |
| Pre-Bid Meeting                  | : 14 <sup>th</sup> November, 2019 at 03:00 PM  |
| Earlier Last Date of Submission  | : 30 <sup>th</sup> January, 2020 at 03:00 PM   |
| Extended Last Date of Submission | : 17 <sup>th</sup> February, 2020 at 03:00 PM  |
| Bid opening                      | : 18 <sup>th</sup> February, 2020 at 03:15 P.M |

**The following revised and additional specification will be added:-**

- 1. Page No. 10, In Technical Specifications, Point No. 1, “Sub Point: j”:**  
**For**  
Colour Display  
**Read**  
Console Should be compatible to Bone Mill
- 2. Page No. 10, In Technical Specifications, Point No. 3, “Sub Point: a”:**  
**For:**  
Maximum Speed 50000 to 75000 RPM  
**Read:**  
Maximum Speed 50000 to 60000 RPM
- 3. Page No. 10, In Technical Specifications, Point No. 9:**  
**For:**  
Pin Collet for Universal Driver  
**Read:**  
Deleted