

Date: - 22<sup>nd</sup>February 2016

### Corrigendum

### for

### Equipments required for Department of Orthopaedics

NIT Issue Date	:	21 <sup>st</sup> January 2016.
NIT No.	:	Admn/Tender/Orthopaedics/2015-AIIMS.JDH
Pre Bid Meeting held on	:	09 <sup>th</sup> February,2016 at 11:00 AM
Last Date of Submission	:	25 <sup>th</sup> February, 2016 at 03:00 PM
Revised Last Date of Submission	:	07 <sup>th</sup> March, 2016 at 03:00 PM

### 1. The following revised and additional specification will be added:-

## 1. Page No. 11, S.No. 1, Bone Forceps Set, Technical Specification of Bone Forceps Set, Point No. 06:

For

Company should have European CE certificates and USFDA certificates of International standard.

#### Read

Company should have European CE certificates or USFDA certificates of International standard and should have an average annual average turnover in the excution of similer work, equal to or in excess of 50 crores per annum for the past 3 years.

### 2. Page No. 12, S.No. 2, Large Fragment Set, Technical Specification of Large Fragment Set, Point No. 06:

For

Company should have European CE certificates and USFDA certificates of International standard.

### Read

Company should have European CE certificates or USFDA certificates of International standard and should have an average annual average turnover in the excution of similer work, equal to or in excess of 50 crores per annum for the past 3 years.

# 3. Page No. 13, S.No. 3, Small Fragment Set, Technical Specification of Small Fragment Set, Point No. 06:

#### For

Company should have European CE certificates and USFDA certificates of International standard.

### Read

Company should have European CE certificates or USFDA certificates of International standard and should have an average annual average turnover in the excution of similer work, equal to or in excess of 50 crores per annum for the past 3 years.

## 4. Page No. 13, S.No. 4, Wire instruments Set, Technical Specification of Wire instruments, Point No. 06:

#### For

Company should have European CE certificates and USFDA certificates of International standard.

### Read

Company should have European CE certificates or USFDA certificates of International standard and should have an average annual average turnover in the excution of similer work, equal to or in excess of 50 crores per annum for the past 3 years.

#### 5. Page No. 16, S.No. 6, DHS Set, Technical Specification of DHS Set, Point No. 06: For

Company should have European CE certificates and USFDA certificates of International standard.

### Read

Company should have European CE certificates or USFDA certificates of International standard and should have an average annual average turnover in the excution of similer work, equal to or in excess of 50 crores per annum for the past 3 years.

6. Page No. 22, S.No. 11, General Orthopaedic Instruments for Arthroplasty, Technical Specification of General Orthopaedic Instruments for Arthroplasty, Point No. 06: For

Company should have European CE certificates and USFDA certificates of International standard.

### Read

Company should have European CE certificates or USFDA certificates of International standard and should have an average annual average turnover in the excution of similer work, equal to or in excess of 50 crores per annum for the past 3 years.

### 7. Page No. 25, S.No. 13, DVT Pumps/ Intermittent Pneumatic Compression Device for Orthopaedics, Point No. 05:

### For

Controller should have USB port to make software updates easy.

### Read

Controller should preferably have USB port to make software updates easy.

### 8. Page No. 25, S.No. 13, DVT Pumps/ Intermittent Pneumatic Compression Device for Orthopaedics, Point No. 09:

For

Battery backup for 8 hours for uninterrupted compression.

Read

Battery backup for minimum 6 hours or more for uninterrupted compression.

- 9. Page No. 25, S.No. 13, DVT Pumps/ Intermittent Pneumatic Compression Device for Orthopaedics, after Point No. 12: Added Point No. 13: Should be supplied with - Knee Length medium Sleeves - 25 pairs. , Thigh length medium sleeves - 25 pairs.
- 10. Page No. 25, S.No. 13, DVT Pumps/ Intermittent Pneumatic Compression Device for Orthopaedics, after Point No. 13: Added Point No. 14:

All accessories to be quoted separately.

Administrative Officer AIIMS, Jodhpur