

Invitation of quotation  
for  
Supply of Aneurysm Clips for the Department of  
Neurosurgery

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/01-04/2018-AIIMS.JDH

Inquiry Issue Date : 06<sup>th</sup> April, 2018

Last Date of Submission : 12<sup>th</sup> April, 2018 at 05:00 PM.



**All India Institute of Medical Sciences, Jodhpur**

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: [procurement@aiimsjodhpur.edu.in](mailto:procurement@aiimsjodhpur.edu.in)

[www.aiimsjodhpur.edu.in](http://www.aiimsjodhpur.edu.in)

**Invitation of quotation for Supply of Aneurysm Clips for the  
Department of Neurosurgery at AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Aneurysm Clips for the Department of Neurosurgery for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 12.04.2018 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR SUPPLY OF ANEURYSM CLIPS FOR THE  
DEPARTMENT OF NEUROSURGERY AGAINST INQUIRY NO.  
ADMN/GEN/01-04/2018-AIIMS.JDH” DUE ON 12.04.2018 05.00 PM”**

**1. Terms & Conditions:**

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
  - Firm shall be registered with the Government of Rajasthan / Central Government.
  - The firm shall have valid GST/VAT/CST/ST/Other taxes and IT PAN.
  - **The firm should not be black listed by any Govt. Agency/Dept.**

## **INQUIRY NO. Admin/Gen/01-04/2018-AIIMS.JDH**

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

### **2. Special Terms & Conditions:**

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**
- D) **Inspection committee will check the products thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then AIIMS, Jodhpur has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained**

**Administrative Officer**

Encl.: Annexure 1 (Specification)  
Annexure 2 (Format of price bid)

**Annexure 1**

**Technical specifications of Aneurysm clips:**

The instruments should be light weight, made up of non-magnetic Titanium/Cobalt alloy with ISO 9001 / CE / PIS or equivalent rating. The dimensions of instruments may vary 5-10% from specification mentioned.

All clips should be manufactured by same manufacturer and should be compatible with the Yasargil Clip Applicator available at AIIMS Jodhpur.

It should be CE approved.

**Aneurysm Clips:**

- The aneurysm clip should be made of Titanium/Cobalt alloy.
- The clip should be MRI safe.
- The clip should have box lock joint to avoid scissoring affect.
- The blade profile should have pyramid like impression. It should be atraumatic.
- The clip should ensure even pressure on the complete blade surface.
- Each clip should have individual serial no. to ensure complete tracking of manufacturing process, material, closing force, sterilization process.
- The clip should be supplied in sterile pack.
- The Package of clip should mention article no, expiry date, closing force.
- The clip should be color coded.
- The manufacturer should quote rates for clips (all sizes) for next 5 years.

<b>S. No.</b>	<b>Specifications of Aneurysm Clips</b>	<b>Blade Length</b>	<b>Qty</b>
1.	Yasargil Permanent Standard Aneurysm clip Straight	7mm	3
2.	Yasargil Permanent Standard Aneurysm clip Straight	9mm	3
3.	Yasargil Permanent Standard Aneurysm clip Straight	11mm	1
4.	Yasargil Permanent Standard Aneurysm Clip Curved	6.5mm	3
5.	Yasargil Permanent Standard Aneurysm Clip Curved	8.3mm	3
6.	Yasargil Permanent Standard Aneurysm Clip Curved	10.2mm	1
7.	Yasargil Permanent Standard Aneurysm Clip Straight with 3.5mm Fenestration	3/8.1mm	1
8.	Yasargil Permanent Standard Aneurysm Clip Straight with 3.5mm Fenestration	4/9.1mm	1
9.	Yasargil Permanent Standard Aneurysm Clip Straight with 3.5mm Fenestration	5/10.1mm	1
10.	Yasargil Permanent Standard Aneurysm Clip Straight with 5.0mm Fenestration	3/9.8mm	1
11.	Yasargil Temporary Standard Aneurysm clip Curved	8.3	2
12.	Yasargil Temporary Standard Aneurysm clip Straight	9mm	2

**Note: - The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier**

**[On the letterhead of firm]**

**ANNEXURE "2"**  
**PRICE BIDFORM**

To,

Administrative Officer,  
AIIMS, Jodhpur.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF ANEURYSM CLIPS AND APPLICATOR FOR THE DEPARTMENT OF NEUROSURGERY AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/01-04/2018-AIIMS.JDH" DUE ON 12.04.2018 05.00 PM for Supply of Aneurysm Clips and Applicator for the Department of Neurosurgery at AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No	Specifications of Aneurysm Clips	Blade Length	Qty	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Other taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1	Yasargil Permanent Standard Aneurysm clip Straight	7mm	3						
2	Yasargil Permanent Standard Aneurysm clip Straight	9mm	3						
3	Yasargil Permanent Standard Aneurysm clip Straight	11mm	1						
4	Yasargil Permanent Standard Aneurysm Clip Curved	6.5mm	3						
5	Yasargil Permanent Standard Aneurysm Clip Curved	8.3mm	3						
6	Yasargil Permanent Standard Aneurysm Clip Curved	10.2m m	1						
7	Yasargil Permanent Standard Aneurysm Clip Straight with 3.5mm Fenestration	3/8.1m m	1						
8	Yasargil Permanent Standard Aneurysm Clip Straight with 3.5mm Fenestration	4/9.1m m	1						
9	Yasargil Permanent Standard Aneurysm Clip Straight with 3.5mm Fenestration	5/10.1 mm	1						
10	Yasargil Permanent Standard Aneurysm Clip Straight with 5.0mm Fenestration	3/9.8m m	1						

**INQUIRY NO. Admin/Gen/01-04/2018-AIIMS.JDH**

11	Yasargil Temporary Standard Aneurysm clip Curved	8.3	2						
12	Yasargil Temporary Standard Aneurysm clip Straight	9mm	2						
<b>Total Amount In Figure</b> (Inclusive of all Taxes and Other Charges)									
<b>Total Amount In Word</b> (Inclusive of all Taxes and Other Charges)									

**Note:**

1. L1 will decided on composite basis.
2. The Bidder must quote only single Make & Model.
3. The bidder must quoted their quotation only in above said format.
4. The bidder may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date \_\_\_\_\_

Place \_\_\_\_\_

(Signature of Authorized Person) \_\_\_\_\_

(Name) \_\_\_\_\_

Name of Firm/Company/Agency \_\_\_\_\_

Phone No. \_\_\_\_\_

Email: \_\_\_\_\_