

Invitation of quotation  
for  
Supply of Pediatric Spirometer  
At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/03-08/2018-AIIMS.JDH

Inquiry Issue Date : 21<sup>st</sup> December, 2018

Last Date of Submission : 27<sup>th</sup> December, 2018 at 05:00 PM.



**All India Institute of Medical Sciences, Jodhpur**

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: [procurement@aiimsjodhpur.edu.in](mailto:procurement@aiimsjodhpur.edu.in)

[www.aiimsjodhpur.edu.in](http://www.aiimsjodhpur.edu.in)

**Invitation of quotation for Supply of Pediatric Spirometer at  
AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Pediatric Spirometer for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 27.12.2018 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR SUPPLY OF PEDIATRIC SPIROMETER  
AGAINST INQUIRY NO. ADMN/GEN/03-08/2018-AIIMS.JDH” DUE ON  
27.12.2018 05.00 PM”**

**1. Terms & Conditions:**

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
  - The firm shall have valid GST/Other taxes and IT PAN.
  - **The firm should not be black listed by any Govt. Agency/Dept.**

**INQUIRY NO. Admin/Gen/03-08/2018-AIIMS.JDH**

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

**2. Special Terms & Conditions:**

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The Bidder must submit the GSTIN Registration and PAN Card self-attested copy with the quotation.**
- D) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

**Administrative Officer**

Encl.: Annexure 1 (Specification)  
Annexure 2 (Format of price bid)

**Annexure 1**

S. No	Particular	Specifications	Qty.
1.	<b>Pediatric Spirometer</b>	<p><b><u>Specifications Pediatric Spirometer:-</u></b>                      The Unit Should be light weight, exclusively designed to carry out examination of pulmonary system and to measure &amp; display lung function parameters along with tidal breath analysis for babies, Capnometry and Emphysema diagnostic, operable on power from the Laptop USB Port, not requiring large installation space (Compact). It should come with an incentive display for cooperation free lung testing for preschool children.                      It should be designed for the following measurement protocols</p> <ul style="list-style-type: none"> <li>• Spirometer with Pre/post comparison, flow/Vol and Vol/Time loops/Curves and Trend Loops.</li> <li>• The Tests should be as per ATS/ERS combined criterion and the report should clearly show compliance to these.</li> <li>• It should also have facility for tidal breath analysis for small babies 2 Weeks and older.</li> <li>• Individual interpretation assistance.</li> <li>• Patient Co-operation Display.</li> </ul> <p>The System should incorporate a light weight, precision Ultrasonic sensor free from any kind of frictional inefficiencies which is absolutely insensitive to moisture.</p> <p>It should have an open kind of breathing system for preventing cross contamination with the following measurement ranges:-</p> <p><b><u>Flow Measurement</u></b>                      Range: 0 to +/- 20 Ltr/Sec with resolution of 10 ml</p> <p><b><u>Volume measurement</u></b>                      Range: 0 to 20 ltrs with resolution of 10 ml</p> <p><b><u>Corrections</u></b>                      Inspiratory Gas Qty: BTPS automatic breathe by breath</p> <p><b><u>The Machine should measure the following parameters:</u></b></p> <p>VC in VC ex, VC max, IRV, ERV, VT, IC, Frequency, MVV, ti, te,ti/te,FVC,FEV1, FEV,FEV 0.5,FEV 1/VC max, FEV 1/FVC ex, MEF25, MEF25-75,PEF,PIF,AREA Under expiratory f/v curve and tidal breath analysis, Capnometry and Emphysema diagnostic.</p> <p>The Unit Should be supplied complete with hardware and software including a Laptop and a suitable printer from a respected brand like Lenovo, Son, Dell or HP.</p>	01 Nos.

**Note:-The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

**[On the letterhead of firm]**

**ANNEXURE "2"**  
**PRICE BIDFORM**

To,

Administrative Officer,  
AIIMS, Jodhpur.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF PEDIATRIC SPIROMETER AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/03-08/2018-AIIMS.JDH" DUE ON 27.12.2018 05.00 PM for Supply of Pediatric Spirometer at AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty	Quote Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	MRP
1.	Pediatric Spirometer <b><u>Specifications:-</u></b> As per annexure – 1.	01 Nos.					

**Note:-**

- The Bidder must quote only single Make & Model.**
- The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be **REJECTED**.
- Catalog must be attached with quotation for technical evaluation.
- The Bidder must submit the GSTIN Registration and PAN Card self-attested copy with the quotation.
- The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

**Date**\_\_\_\_\_

**Place**\_\_\_\_\_

**(Signature of Authorized Person)**\_\_\_\_\_

**(Name)**\_\_\_\_\_

**Name of Firm/Company/Agency**\_\_\_\_\_

**Phone No.**\_\_\_\_\_

**Email:**\_\_\_\_\_