

Invitation of quotation  
for  
Supply of Binocular Microscope  
At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/09-02/2021-AIIMS.JDH

Inquiry Issue Date : 08<sup>th</sup> April, 2021

Last Date of Submission : 15<sup>th</sup> April, 2021 at 03:00 PM.



**All India Institute of Medical Sciences, Jodhpur**

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: [procurement@aiimsjodhpur.edu.in](mailto:procurement@aiimsjodhpur.edu.in)

[www.aiimsjodhpur.edu.in](http://www.aiimsjodhpur.edu.in)

**Invitation of quotation for Supply of Binocular Microscope at  
AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Binocular Microscope for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 15.04.2021 03:00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR SUPPLY OF BINOCULAR MICROSCOPE  
AGAINST INQUIRY NO. ADMN/GEN/09-02/2021-AIIMS.JDH” DUE ON  
15.04.2021 03:00 PM”**

**1. Terms & Conditions:**

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
  - The firm shall have valid GST No.
  - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior

**INQUIRY NO. Admin/Gen/09-02/2021-AIIMS.JDH**

confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

**2. Special Terms & Conditions:**

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

**Administrative Officer**

Encl.: Annexure 1 (Specification)  
Annexure 2 (Format of price bid)

**Annexure 1**

S. No.	Particular	Specification	Required Qty.																														
1.	Binocular Microscope	<p><b><u>The equipment should have the following features:-</u></b></p> <table border="1"> <tr> <td data-bbox="363 398 555 472">Optical System</td> <td data-bbox="560 398 587 472">:</td> <td data-bbox="592 398 1326 472">UIS2 (Universal Infinity Corrected) Optical System with anti-fungal coating</td> </tr> <tr> <td data-bbox="363 479 555 577">Focusing system</td> <td data-bbox="560 479 587 577">:</td> <td data-bbox="592 479 1326 577">Co-axil coarse and fine controls with a focus adjustment and fine adjustment knobs. Coarse focus range 20mm. Fine focus rotation 0.25mm.</td> </tr> <tr> <td data-bbox="363 584 555 728">Body</td> <td data-bbox="560 584 587 728">:</td> <td data-bbox="592 584 1326 728">Aluminium die-cast body with critical movement based on call bearing and wire guides thereby ensuring smooth and precise manipulation, optics should be anti-fungus treated for component durability</td> </tr> <tr> <td data-bbox="363 734 555 878">Observation type</td> <td data-bbox="560 734 587 878">:</td> <td data-bbox="592 734 1326 878">Siedentopf designed 30° inclined binocular observation tube which should support filed of view of 20m or more. It should have a inter-pupillary distance adjustment from 48~75mm.</td> </tr> <tr> <td data-bbox="363 884 555 987">Objective</td> <td data-bbox="560 884 587 987">:</td> <td data-bbox="592 884 1326 987">Plan Achromat objective of 4x/0.10, 10x/0.25, 40x/0.65 (spring) and 100x/1.25 (spring, oil) should be quoted.</td> </tr> <tr> <td data-bbox="363 994 555 1097">Nosepiece</td> <td data-bbox="560 994 587 1097">:</td> <td data-bbox="592 994 1326 1097">Quintuple revolving nosepiece with inward tilt (facing away from the operator) &amp; precision ball-bearing mechanism with positive click stop.</td> </tr> <tr> <td data-bbox="363 1104 555 1247">Condenser</td> <td data-bbox="560 1104 587 1247">:</td> <td data-bbox="592 1104 1326 1247">Abbe Condenser with aperture iris diaphragm having NA 1.25 focusable with rack * pinion and a continuously variable iris diaphragm &amp; rack and pinon mounted condenser holder.</td> </tr> <tr> <td data-bbox="363 1254 555 1357">Illumination</td> <td data-bbox="560 1254 587 1357">:</td> <td data-bbox="592 1254 1326 1357">3W LED light source high brightness, long life (30,000 hrs) and SMPS circuit for constant voltage output.</td> </tr> <tr> <td data-bbox="363 1364 555 1507">Mechanical Stage</td> <td data-bbox="560 1364 587 1507">:</td> <td data-bbox="592 1364 1326 1507">Rectangular Rco-axil low drive mechanical strage with wire movement (120mmx132mm) (+/-5mm) with traverse area of 30mm x 76 (+/-5mm)with single slide holder.</td> </tr> <tr> <td data-bbox="363 1514 555 1617">Eyepiece</td> <td data-bbox="560 1514 587 1617">:</td> <td data-bbox="592 1514 1326 1617">Paired wide-field eyepiece of 10x having field of view of 20mm or higher with diopter setting of +/- should be available on one eyepiece position.</td> </tr> </table>	Optical System	:	UIS2 (Universal Infinity Corrected) Optical System with anti-fungal coating	Focusing system	:	Co-axil coarse and fine controls with a focus adjustment and fine adjustment knobs. Coarse focus range 20mm. Fine focus rotation 0.25mm.	Body	:	Aluminium die-cast body with critical movement based on call bearing and wire guides thereby ensuring smooth and precise manipulation, optics should be anti-fungus treated for component durability	Observation type	:	Siedentopf designed 30° inclined binocular observation tube which should support filed of view of 20m or more. It should have a inter-pupillary distance adjustment from 48~75mm.	Objective	:	Plan Achromat objective of 4x/0.10, 10x/0.25, 40x/0.65 (spring) and 100x/1.25 (spring, oil) should be quoted.	Nosepiece	:	Quintuple revolving nosepiece with inward tilt (facing away from the operator) & precision ball-bearing mechanism with positive click stop.	Condenser	:	Abbe Condenser with aperture iris diaphragm having NA 1.25 focusable with rack * pinion and a continuously variable iris diaphragm & rack and pinon mounted condenser holder.	Illumination	:	3W LED light source high brightness, long life (30,000 hrs) and SMPS circuit for constant voltage output.	Mechanical Stage	:	Rectangular Rco-axil low drive mechanical strage with wire movement (120mmx132mm) (+/-5mm) with traverse area of 30mm x 76 (+/-5mm)with single slide holder.	Eyepiece	:	Paired wide-field eyepiece of 10x having field of view of 20mm or higher with diopter setting of +/- should be available on one eyepiece position.	03 Nos.
Optical System	:	UIS2 (Universal Infinity Corrected) Optical System with anti-fungal coating																															
Focusing system	:	Co-axil coarse and fine controls with a focus adjustment and fine adjustment knobs. Coarse focus range 20mm. Fine focus rotation 0.25mm.																															
Body	:	Aluminium die-cast body with critical movement based on call bearing and wire guides thereby ensuring smooth and precise manipulation, optics should be anti-fungus treated for component durability																															
Observation type	:	Siedentopf designed 30° inclined binocular observation tube which should support filed of view of 20m or more. It should have a inter-pupillary distance adjustment from 48~75mm.																															
Objective	:	Plan Achromat objective of 4x/0.10, 10x/0.25, 40x/0.65 (spring) and 100x/1.25 (spring, oil) should be quoted.																															
Nosepiece	:	Quintuple revolving nosepiece with inward tilt (facing away from the operator) & precision ball-bearing mechanism with positive click stop.																															
Condenser	:	Abbe Condenser with aperture iris diaphragm having NA 1.25 focusable with rack * pinion and a continuously variable iris diaphragm & rack and pinon mounted condenser holder.																															
Illumination	:	3W LED light source high brightness, long life (30,000 hrs) and SMPS circuit for constant voltage output.																															
Mechanical Stage	:	Rectangular Rco-axil low drive mechanical strage with wire movement (120mmx132mm) (+/-5mm) with traverse area of 30mm x 76 (+/-5mm)with single slide holder.																															
Eyepiece	:	Paired wide-field eyepiece of 10x having field of view of 20mm or higher with diopter setting of +/- should be available on one eyepiece position.																															

**Note: -**

- **Catalog/ Brochure must be attached for Technical Evaluation.**
- **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

**[On the letterhead of firm]**

**ANNEXURE "2"**  
**PRICE BIDFORM**

To,  
Administrative Officer,  
AIIMS, Jodhpur.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF BINOCULAR MICROSCOPE AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/09-02/2021-AIIMS.JDH" DUE ON 15.04.2021 03:00 PM for Supply of Binocular Microscope at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Binocular Microscope <b>Specification:-</b> As per Annexure-1	03 Nos.						

**Note:-**

1. The Bidder must quote only single Make & Model.
2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date \_\_\_\_\_

(Name) \_\_\_\_\_

Place \_\_\_\_\_

Name of Firm/Company/Agency \_\_\_\_\_

GSTIN No.: \_\_\_\_\_

Bank Name:- \_\_\_\_\_

Bank Account No.: \_\_\_\_\_

IFSC Code:- \_\_\_\_\_

Branch Name: \_\_\_\_\_

Phone No. \_\_\_\_\_

Email: \_\_\_\_\_

(Signature of Authorized Person) \_\_\_\_\_

Seal: \_\_\_\_\_