

Invitation of quotation
for
Supply of Ophthalmoscope LED with USB
Rechargeable Handle

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/11-04/2021-AIIMS.JDH

Inquiry Issue Date : 28th October, 2021

Last Date of Submission : 08th November, 2021 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Supply of Ophthalmoscope LED with USB Rechargeable Handle at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Ophthalmoscope LED with USB Rechargeable Handle for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 08.11.2021 03:00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR SUPPLY OF OPHTHALMOSCOPE LED WITH USB RECHARGEABLE HANDLE AGAINST INQUIRY NO. ADMN/GEN/11-04/2021-AIIMS.JDH” DUE ON 08.11.2021 03:00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST No.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior

INQUIRY NO. Admin/Gen/11-04/2021-AIIMS.JDH

confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Deputy Director (Administration)

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

S. No.	Particular	Qty
1.	Ophthalmoscope LED with USB Rechargeable Handle	01 Nos

Specifications of Ophthalmoscope LED with USB Rechargeable Handle

The Ophthalmoscope with superior aspherical optics.

The separation of illumination and observation beam by means of the aspherical optical system avoids corneal and iris reflex. Ophthalmoscope should allow you to view the whole of the illuminated section of the retina even with a small pupil.

Unique optical system.	*	Corneal and iris reflex is eliminated. A full view is guarantee even with small pupils.																		
LED control.	*	LED in high quality with continuous brightness																		
Recessed, multi-coated viewing Window	*	avoids stray light																		
Dustproof housing	*	maintenance-free																		
Optics mounted in precision Metal chassis	*	durable, precise																		
Ergonomic shape	*	instrument fits the orbita comfortably in any position.																		
Soft orbital support	*	protects the users' glasses and steadies the Instrument																		
6 apertures filter, large spot, small spot, hemispot apertures with separated red-free filter	*	Slit, fixation star with polar coordinates, cobalt Blue suitable for large or small pupils. Incorporates slit aperture, fixation star, cobalt blue filter and red-free filter for improved contrast.																		
Lens range :	*	gives good results even 2with high refractive Errors.																		
<table> <tbody> <tr> <td>+ in 1 D steps 1-10</td> <td> </td> <td>15</td> <td> </td> <td>20</td> <td> </td> <td>40</td> <td> </td> <td>D</td> </tr> <tr> <td>- in 1 D steps 1-10</td> <td> </td> <td>15</td> <td> </td> <td>20</td> <td> </td> <td>25</td> <td> </td> <td>35D</td> </tr> </tbody> </table>			+ in 1 D steps 1-10		15		20		40		D	- in 1 D steps 1-10		15		20		25		35D
+ in 1 D steps 1-10		15		20		40		D												
- in 1 D steps 1-10		15		20		25		35D												

- **High Quality Metal construction** for strength and durability.
- **Exclusive Automatic Connector with locking design.** No metal on metal friction which can wear out over time, for a secure fit for the lifetime of your instruments.
- **Deep Battery Discharge Protection:** Integrated safety system protects your investment by preventing batteries from completely draining, thus ensuring maximum battery life.
- **Familiar brightness control:** Rheostat for continuous brightness control with power level indicator.
- **Li-Ion technology:** no "memory effect". Charge your handle how you like without worrying about charge level or battery condition.
- **Quick Charge.**
- Slightly knurled finish of the handle provides a **non-slip grip but is easy to clean and disinfect.**
- **Charge status indicator** in the bottom insert; the indicator shines orange if the battery needs to be recharged. In the BETA 4 USB Handle the indicator blinks green during the charging process and shines permanently green as soon as it is fully charged.
- USB rechargeable Handle with Li-Ion battery with charge status indicator & USB Charger

Note: -

- ❖ **Catalog/ Brochure must be attached for Technical Evaluation.**

INQUIRY NO. Admin/Gen/11-04/2021-AIIMS.JDH

- ❖ The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,
Deputy Director (Administration),
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF OPHTHALMOSCOPE LED WITH USB RECHARGEABLE HANDLE AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/11-04/2021-AIIMS.JDH" DUE ON 08.11.2021 03:00 PM for Supply of Ophthalmoscope LED with USB Rechargeable Handle at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Ophthalmoscope LED with USB Rechargeable Handle Specification:- As per annexure – 1	01 Nos.						

Note:-

1. The Bidder must quote only single Make & Model.
2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

Place _____

(Name) _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____