

**Invitation of quotation**  
**for**  
**Supply & Installation of 5.5 Kv Online UPS for**  
**High End Ophthalmic Machinery for Department**  
**of Ophthalmology**  
**At**  
**All India Institute of Medical Sciences, Jodhpur**

Inquiry No.: : Admin/Gen/17-20/2017-AIIMS.JDH

Inquiry Issue Date : 08<sup>th</sup> November, 2017

Last Date of Submission : 14<sup>th</sup> November, 2017 at 05:00 PM.



**All India Institute of Medical Sciences, Jodhpur**

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: [procurement@aiimsjodhpur.edu.in](mailto:procurement@aiimsjodhpur.edu.in)

[www.aiimsjodhpur.edu.in](http://www.aiimsjodhpur.edu.in)

**Invitation of quotation for Supply & Installation of 5.5 Kv Online UPS for High End Ophthalmic Machinery for Department of Ophthalmology at AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply & Installation of 5.5 Kv Online UPS for High End Ophthalmic Machinery for Department of Ophthalmology for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 14.11.2017 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR SUPPLY & INSTALLATION OF 5.5 KV ONLINE UPS FOR HIGH END OPHTHALMIC MACHINERY FOR DEPARTMENT OF OPHTHALMOLOGY AGAINST INQUIRY NO. ADMN/GEN/17-20/2017-AIIMS.JDH” DUE ON 14.11.2017 05.00 PM”**

**1. Terms & Conditions:**

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
  - Firm shall be registered with the Government of Rajasthan / Central

**INQUIRY NO. Admin/Gen/17-20/2017-AIIMS.JDH**

Government.

- The firm shall have valid GST/VAT/CST/ST/Other taxes and IT PAN.
- **The firm should not be black listed by any Govt. Agency/Dept.**

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

**2. Special Terms & Conditions:**

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**
- D) **Inspection committee will check the products thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then AIIMS, Jodhpur has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained**

**Administrative Officer**

Encl.: Annexure 1 (Specification)  
Annexure 2 (Format of price bid)

**Annexure 1**

**TECHNICAL SPECIFICATIONS OF ONLINE UPS**

Rated Capacity	5.5 KVA /96 V
Output Power	4400 Watts
Battery	8 Battery System (96 VDC)
Waveform	Pure Sinewave
Warranty	5 Years

**Key Features**

- Advanced Battery Management (ABM) for Longer Battery Life and Quick Charging
- DSP based PWM Technology using Hi-Grade IGBT'S
- LCD Panel Display for status and faults
- Static By Pass Switch for fast Switch Over (Optional)
- Generator Compatible
- In-built TDR for Compressor Based Application, e.g. Air Conditioners

**Display**

- MAINS ON
- UPS ON
- BATTERY CHARGING
- OVERLOADED
- BATTERY LOW
- FAULT INDICATION

**Technical Specifications**

**Input**

Input Voltage (Standard Range) 100 V ~ 295 V

Input Voltage (Narrow Range) 180 V ~ 265 V

**Output**

Output Voltage (Mains Mode) Same as Input

Output Voltage (UPS Mode) 220V ± 10%

Output Waveform (Mains Mode) Same as Input

Output Waveform (UPS Mode) Pure Sinewave

**Changeover**

Switchover from Mains to UPS and UPS to Mains Automatic

UPS Transfer Time ≤15 msec.

**Technology**

UPS Overload / UPS Short-Circuit 110% / 250%

Browns on Mains Voltage 100V ± 40V

Efficiency (Battery Mode) >82%

Auto Reset Feature Yes

Input DC Range 80 ~ 114 V ± 1 Volt

Charging Current 10 Amps ± 5 Amps

Auto Reset Feature Yes

Separate Output with TDR 3 ~ 5 minutes TDR

Operating Temperature 0 to 45° C

**Note: The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

**[On the letterhead of firm]**

**ANNEXURE "2"**  
**PRICE BIDFORM**

To,

Administrative Officer,  
AIIMS, Jodhpur.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY & INSTALLATION OF 5.5 KV ONLINE UPS FOR HIGH END OPHTHALMIC MACHINERY FOR DEPARTMENT OF OPHTHALMOLOGY AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/17-20/2017-AIIMS.JDH" DUE ON 14.11.2017 05.00 PM for Supply & Installation of 5.5 Kv Online UPS for High End Ophthalmic Machinery for Department of Ophthalmology at AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Installation/Other Taxes	Price/Unit Inclusive of GST (INR)	Total Cost Inclusive of GST/Installation/Other Taxes	MRP
1.	Supply & Installation of 5.5 Kv Online UPS for High End Ophthalmic Machinery for Department of Ophthalmology	02 Nos						

**Note:**

- The Bidder must quote only single Make & Model.**
- The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Date \_\_\_\_\_

Place \_\_\_\_\_

(Signature of Authorized Person) \_\_\_\_\_

(Name) \_\_\_\_\_

Name of Firm/Company/Agency \_\_\_\_\_

Phone No. \_\_\_\_\_

Email: \_\_\_\_\_