

Invitation of quotation
for
Eternal Feeding Pump
At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/26-01(ii)/2018-AIIMS.JDH

Inquiry Issue Date : 01st May, 2018

Last Date of Submission : 07th May, 2018 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

**Invitation of quotation for Eternal Feeding Pump at AIIMS
Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Eternal Feeding Pump for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 07.05.2018 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR ETERNAL FEEDING PUMP AGAINST INQUIRY
NO. ADMN/GEN/26-01(ii)/2018-AIIMS.JDH” DUE ON 07.05.2018 05.00
PM”**

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/VAT/CST/ST/Other taxes and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**

INQUIRY NO. Admin/Gen/26-01(ii)/2018-AIIMS.JDH

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**
- D) **Inspection committee will check the products thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then AIIMS, Jodhpur has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained.**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

Technical Specification

S. No.	Particular	Specification	Qty.																																																	
1.	Eternal Feeding Pump	<table border="1"> <tr> <td data-bbox="370 421 651 456">FLOW RATE</td> <td data-bbox="655 421 1386 456">1-100ml/h (1ml/h increments)</td> </tr> <tr> <td data-bbox="370 463 651 499">CHANGE</td> <td data-bbox="655 463 1386 499">100-600ml/h (5ml/h increments)</td> </tr> <tr> <td data-bbox="370 506 651 542">FLOW RATE</td> <td data-bbox="655 506 1386 542">$\pm 7\%$ at 50ml/h</td> </tr> <tr> <td data-bbox="370 548 651 584">ACCURACY</td> <td data-bbox="655 548 1386 584"></td> </tr> <tr> <td data-bbox="370 591 651 627">FEEDING</td> <td data-bbox="655 591 1386 627">1-100ml/h (1ml/h increments)</td> </tr> <tr> <td data-bbox="370 633 651 669">VOLUME</td> <td data-bbox="655 633 1386 669">100-500ml/h (5ml/h increments)</td> </tr> <tr> <td data-bbox="370 676 651 712">FEEDING MODE</td> <td data-bbox="655 676 1386 712">Continuous</td> </tr> <tr> <td data-bbox="370 719 651 754">PRIMING</td> <td data-bbox="655 719 1386 754">Automatic & manual priming availability</td> </tr> <tr> <td data-bbox="370 761 651 797">COUNTER</td> <td data-bbox="655 761 1386 797">Preferably cumulative feeding volume Counters from 0.001L to 99.999L</td> </tr> <tr> <td data-bbox="370 804 651 840">DATA EVENT</td> <td data-bbox="655 804 1386 840">Feeding history 24-72hrs or 250 Events (more will be preferable)</td> </tr> <tr> <td data-bbox="370 846 651 882">LOG</td> <td data-bbox="655 846 1386 882"></td> </tr> <tr> <td data-bbox="370 889 651 925">NIGHT MODE</td> <td data-bbox="655 889 1386 925">Night mode decreases brightness of Screen & the power LED (preferable)</td> </tr> <tr> <td data-bbox="370 931 651 967">KEYPAD LOCK</td> <td data-bbox="655 931 1386 967">Possibility to lock the keypad to prevent Unintentional key press</td> </tr> <tr> <td data-bbox="370 974 651 1010">INTENDED USE</td> <td data-bbox="655 974 1386 1010">For external feeding use only compatible to hospital conventional feed.</td> </tr> <tr> <td data-bbox="370 1016 651 1052">INFORMATION</td> <td data-bbox="655 1016 1386 1052">Target volume almost reached battery almost Discharge, start reminder, technical information</td> </tr> <tr> <td data-bbox="370 1059 651 1095">ALARMS</td> <td data-bbox="655 1059 1386 1095">Target volume reached door open, wrong set Installation, downstream occlusion, occlusion, Empty bag/air in line, empty battery, technical Information</td> </tr> <tr> <td data-bbox="370 1102 651 1137">PUMPING</td> <td data-bbox="655 1102 1386 1137">Linear peristaltic pumping system</td> </tr> <tr> <td data-bbox="370 1144 651 1180">MECHANISM</td> <td data-bbox="655 1144 1386 1180"></td> </tr> <tr> <td data-bbox="370 1187 651 1223">DISPLAY</td> <td data-bbox="655 1187 1386 1223">LCD with good visibility</td> </tr> <tr> <td data-bbox="370 1229 651 1265">BATTERY</td> <td data-bbox="655 1229 1386 1265">Full battery charging time: 5-6 hrs. Battery life minimum of 12-15 hrs. Once fully charged</td> </tr> <tr> <td data-bbox="370 1272 651 1308">FLUSH</td> <td data-bbox="655 1272 1386 1308">Programmable flushing capabilities</td> </tr> <tr> <td data-bbox="370 1314 651 1350">FREE FLOW</td> <td data-bbox="655 1314 1386 1350">Preventing the risk of free flow when door is opened</td> </tr> <tr> <td data-bbox="370 1357 651 1393">PROTECTION</td> <td data-bbox="655 1357 1386 1393">When the set is disengaged to ensure the patient Safety and an adequate delivery of nutrition</td> </tr> <tr> <td data-bbox="370 1400 651 1435">SET LENGTH</td> <td data-bbox="655 1400 1386 1435">Adjustable anti free flow clamp positioning</td> </tr> <tr> <td data-bbox="370 1442 651 1478"></td> <td data-bbox="655 1442 1386 1478"> <ul style="list-style-type: none"> • Device must be certified by recognized society for safe use like USFDA/ACE/EC • Minimum of two years' warranty. (Under taking must be given otherwise quotation will be similarly rejected) </td> </tr> </table>	FLOW RATE	1-100ml/h (1ml/h increments)	CHANGE	100-600ml/h (5ml/h increments)	FLOW RATE	$\pm 7\%$ at 50ml/h	ACCURACY		FEEDING	1-100ml/h (1ml/h increments)	VOLUME	100-500ml/h (5ml/h increments)	FEEDING MODE	Continuous	PRIMING	Automatic & manual priming availability	COUNTER	Preferably cumulative feeding volume Counters from 0.001L to 99.999L	DATA EVENT	Feeding history 24-72hrs or 250 Events (more will be preferable)	LOG		NIGHT MODE	Night mode decreases brightness of Screen & the power LED (preferable)	KEYPAD LOCK	Possibility to lock the keypad to prevent Unintentional key press	INTENDED USE	For external feeding use only compatible to hospital conventional feed.	INFORMATION	Target volume almost reached battery almost Discharge, start reminder, technical information	ALARMS	Target volume reached door open, wrong set Installation, downstream occlusion, occlusion, Empty bag/air in line, empty battery, technical Information	PUMPING	Linear peristaltic pumping system	MECHANISM		DISPLAY	LCD with good visibility	BATTERY	Full battery charging time: 5-6 hrs. Battery life minimum of 12-15 hrs. Once fully charged	FLUSH	Programmable flushing capabilities	FREE FLOW	Preventing the risk of free flow when door is opened	PROTECTION	When the set is disengaged to ensure the patient Safety and an adequate delivery of nutrition	SET LENGTH	Adjustable anti free flow clamp positioning		<ul style="list-style-type: none"> • Device must be certified by recognized society for safe use like USFDA/ACE/EC • Minimum of two years' warranty. (Under taking must be given otherwise quotation will be similarly rejected)
		FLOW RATE	1-100ml/h (1ml/h increments)																																																	
		CHANGE	100-600ml/h (5ml/h increments)																																																	
		FLOW RATE	$\pm 7\%$ at 50ml/h																																																	
		ACCURACY																																																		
		FEEDING	1-100ml/h (1ml/h increments)																																																	
		VOLUME	100-500ml/h (5ml/h increments)																																																	
		FEEDING MODE	Continuous																																																	
		PRIMING	Automatic & manual priming availability																																																	
		COUNTER	Preferably cumulative feeding volume Counters from 0.001L to 99.999L																																																	
		DATA EVENT	Feeding history 24-72hrs or 250 Events (more will be preferable)																																																	
		LOG																																																		
		NIGHT MODE	Night mode decreases brightness of Screen & the power LED (preferable)																																																	
		KEYPAD LOCK	Possibility to lock the keypad to prevent Unintentional key press																																																	
		INTENDED USE	For external feeding use only compatible to hospital conventional feed.																																																	
		INFORMATION	Target volume almost reached battery almost Discharge, start reminder, technical information																																																	
		ALARMS	Target volume reached door open, wrong set Installation, downstream occlusion, occlusion, Empty bag/air in line, empty battery, technical Information																																																	
PUMPING	Linear peristaltic pumping system																																																			
MECHANISM																																																				
DISPLAY	LCD with good visibility																																																			
BATTERY	Full battery charging time: 5-6 hrs. Battery life minimum of 12-15 hrs. Once fully charged																																																			
FLUSH	Programmable flushing capabilities																																																			
FREE FLOW	Preventing the risk of free flow when door is opened																																																			
PROTECTION	When the set is disengaged to ensure the patient Safety and an adequate delivery of nutrition																																																			
SET LENGTH	Adjustable anti free flow clamp positioning																																																			
	<ul style="list-style-type: none"> • Device must be certified by recognized society for safe use like USFDA/ACE/EC • Minimum of two years' warranty. (Under taking must be given otherwise quotation will be similarly rejected) 																																																			

 04 Nos. |

Note: - The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. **"QUOTATION FOR ETERNAL FEEDING PUMP AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/26-01(ii)/2018-AIIMS.JDH" DUE ON 07.05.2018 05.00 PM** for Eternal Feeding Pump at AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Other taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Eternal Feeding Pump	04 Nos.						

Note:

- The Bidder must quote only single Make & Model.**
- The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED.**
- The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Date _____

Place _____

(Signature of Authorized Person) _____

(Name) _____

Name of Firm/Company/Agency _____

Phone No. _____

Email: _____