

Invitation of quotation
for
Supply of Surgeon Operating Stool for the
Department of Neurosurgery
At
All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/44-06(ii)/2017-AIIMS.JDH

Inquiry Issue Date : 11th September, 2017

Last Date of Submission : 18th September, 2017 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Supply of Surgeon Operating Stool for the Department of Neurosurgery at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Surgeon Operating Stool for the Department of Neurosurgery for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 18.09.2017 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR SUPPLY OF SURGEON OPERATING STOOL FOR THE DEPARTMENT OF NEUROSURGERY AGAINST INQUIRY NO. ADMN/GEN/44-06(ii)/2017-AIIMS.JDH” DUE ON 18.09.2017 05.00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GSTIN and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**

INQUIRY NO. Admin/Gen/44-06(ii)/2017-AIIMS.JDH

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** –within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **The supplier may be asked to submit a sample of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.**
- C) **Catalog must be attached with quotation for technical evaluation.**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

Specification

S. No	Item	Technical Specification	Qty
01.	Surgeon Operating Stool	<p>Technical Specification for Surgeon Operating Stool</p> <ol style="list-style-type: none"> 1. It should have the provision of hands-free height adjustment by way of foot pedal adjustment with hydraulic pump adjustment which is fitted in column of the chair. 2. The Foot pedal should have up movement when pressing down on pedal and down movement on single press & hold of pedal. 3. The pedal should be placed near the base for comfortable movement. 4. The Chair column should be sturdy with minimum 3 pronged sturdy base covered with ABS ergonomic cowling protector, each of the legs (minimum 3) having imported twin wheel 50mm db castors. 5. The breaking mechanism should be placed on the center leg on the ABS cowling protector with one pedal push system for brake & release. 6. It should have adjustable backrest for up & down movement with ergonomic screw mechanism 7. It should have swivel arm rests which may also be removable and can be locked with screw mechanism adjustment. 8. The chair should have rocker pedal swivel locking & the seat should be saddle Seat design for comfort during long operations. 9. The seat backrest & cover should be covered with high grade material which should be easy to clean & disinfect. 10. The height adjustment from the seat to the floor should be – 19 inches to 27 inches. 11. Should be European CE or US FDA approved. 	02 Nos.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. **"QUOTATION FOR SUPPLY & SURGEON OPERATING STOOL FOR THE DEPARTMENT OF NEUROSURGERY AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/44-06(ii)/2017-AIIMS.JDH" DUE ON 18.09.2017 05.00 PM** for Supply & Surgeon Operating Stool for the Department of Neurosurgery at AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty	Quoted Make	Price/Unit Exclusive of TAX (INR)	GST/ VAT/ CST/ST/ Other taxes	Price/ Unit Inclusive of TAX (INR)	Total Amount Inclusive of TAX (INR)	MRP
1.	Surgeon Operating Stool for the Department of Neurosurgery	02 Nos.						

Note:

- 1. The supplier may be asked to submit a sample of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.**
- 2. The Bidder must be quote only single make & model for each items.**
- 3. Catalog must be attached for quoted make & model with quotation for technical evaluation.**

Date_____

Place_____

(Signature of Authorized Person)_____

(Name)_____

Name of Firm/Company/Agency_____

Phone No._____

Email:_____