

Invitation of quotation  
for  
Automated External Defibrillator trainer for BLS  
course

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admn/Gen/14-01/2017-AIIMS.JDH

Inquiry Issue Date : 12<sup>th</sup> April, 2017

Last Date of Submission : 18<sup>th</sup> April, 2017 at 05:00 PM.



**All India Institute of Medical Sciences, Jodhpur**

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: [procurement@aiimsjodhpur.edu.in](mailto:procurement@aiimsjodhpur.edu.in)

[www.aiimsjodhpur.edu.in](http://www.aiimsjodhpur.edu.in)

**Invitation of quotation for Automated External Defibrillator  
trainer for BLS course at AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for supply of Automated External Defibrillator trainer for BLS course for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 18.04.2017 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR SUPPLY OF AUTOMATED EXTERNAL  
DEFIBRILLATOR TRAINER FOR BLS COURSE AGAINST INQUIRY  
NO. ADMN/GEN/14-01/2017-AIIMS.JDH” DUE ON 18.04.2017 05.00 PM”**

**1. Terms & Conditions:**

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
  - The firm shall have valid VAT/ Sales Tax No. and IT PAN.
  - **The firm should not be black listed by any Govt. Agency/Dept.**

**INQUIRY NO. Admn/Gen/14-01/2017-AIIMS.JDH**

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

**2. Special Terms & Conditions:**

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **The supplier may be asked to submit a sample of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.**

**Administrative Officer**

Encl.: Annexure 1 (Specification)  
Annexure 2 (Format of price bid)

**Annexure 1****Specification**

Sr. No	Particular	Description Of Item	Qty
1	Automated External Defibrillator trainer for BLS course	<ol style="list-style-type: none"> <li>1. The AED should provide clear and confident voice instructions with easy-to-follow directions.</li> <li>2. It should have preconfigured scenarios for easy training.</li> <li>3. It should have easy to use pan with volume control and language button.</li> <li>4. It should have five (5) scenarios for varied training opportunities.</li> <li>5. The training pad can be used at least 50-60 times during CPR/AED classes.</li> <li>6. The pad should be pre-connected.</li> <li>7. The pad should have an embedded 'switch' to sense proper pad attachment to the manikin.</li> <li>8. The adhesive on the pads should be silicone based, coupled with the foam structure, for easy application and removal with no residue left behind.</li> <li>9. It should have two sets of adult pads with case, pre-installed batteries, instructions and carrying case.</li> <li>10. It should be compatible with any CPR manikins.</li> <li>11. It should be capable of training in both semi-automatic and fully-automatic versions.</li> <li>12. It should have inbuilt metronome and counts upwards up to 30 to help the student in practicing chest compressions.</li> <li>13. The start, stop function and change of scenarios should be possible with remote control from a distance of 6 feet.</li> <li>14. The AED kit should include one paediatric pad with case.</li> <li>15. It should carry a warranty for three (3) years.</li> </ol>	04

**[On the letterhead of firm]**

**ANNEXURE "2"**  
**PRICE BIDFORM**

To,

Administrative Officer,  
AIIMS, Jodhpur.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. **"QUOTATION FOR SUPPLY OF AUTOMATED EXTRNAL DEFIBRILLATOR TRAINER FOR BLS COURSE AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/14-01/2017-AIIMS.JDH" DUE ON 18.04.2017 05.00 PM** for supply of Automated External Defibrillator trainer for BLS course at AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No.	Particulars	Quoted Make	Qty	Price/Unit Exclusive of TAX (INR)	TAX %	Price/Unit inclusive of TAX (INR)	Total Amount Inclusive of TAX (INR)
1	Automated External Defibrillator trainer for BLS course		04				

Date\_\_\_\_\_

Place\_\_\_\_\_

(Signature of Authorized Person)\_\_\_\_\_

(Name)\_\_\_\_\_

Name of Firm/Company/Agency\_\_\_\_\_

Phone No.\_\_\_\_\_

Email:\_\_\_\_\_