

Invitation of quotation
for
Tourniquet Machine
At

All India Institute of Medical Sciences, Jodhpur

Inquiry No. : Admin/Gen/52-03/2022-AIIMS.JDH

Inquiry Issue Date : 20th June, 2022

Last Date of Submission : 27th June, 2022 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Tourniquet Machine at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Tourniquet Machine for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach the office of the undersigned on or before 27.06.2022 03.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR TOURNIQUET MACHINE AGAINST INQUIRY NO. ADMIN/GEN/52-03/2022-AIIMS.JDH” DUE ON 27.06.2022 03.00 PM”

Terms & Conditions:

A) The quotations received unsealed and after the deadline will not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer submitted through Fax/Email shall not be considered and no correspondence will be entertained in this matter.**

B) Quotations must be in the prescribed proforma (Annexure-2) on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in **“Quotation Box”** located in Administration Block of AIIMS, Jodhpur.

C) Rates must be quoted in **Indian Rupees (INR)** and as per the format specified taxes extra if any must be written separately.

D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)

E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.

F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.

G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.

H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.

I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:

- Firm shall be registered with the Government of Rajasthan / Central Government.
- The firm shall have valid GST/Other taxes and IT PAN.

The firm should not be black listed by any Govt. Agency/Dept

A) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

INQUIRY NO. Admin/Gen/52-03/2022-AIIMS.JDH

- B) **Delivery Period** – within 30 days from Purchase order.
- C) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- D) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- E) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- F) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- G) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

1. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) The supplier may be asked for demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur. The expenditure incurred for demonstrating the items will be borne by the supplier.

Deputy Director (Admin)

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

S.No.	Particulars	Qty	Recommended Make
1.	Tourniquet Machine with Bilateral Cuff for Upper and Lower Limbs (2 Each)	1	Medtronic/Zimmer/Depuy Synthes

SPECIFICATION OF TOURNIQUET MACHINE

- Cuff pressure range 10 to 430 mmHg Pressure regulation \pm 10mm hg of set point
- Online setting Increase and decrease in pressuresettings on jump of ten
- Timer Range from 9 hrs to 59 minutes,decrement mode from set value.
- Internal least count One minute , Internal 1/1000second
- Memory Pressure set in earlier session is stored and displayed, when machine is switched ON again.
- Pressure Regulation Maintains set pressure under leakconditions.
- Backup 3 hours battery backup (on full charge) Switchover Automatic switchover from mains to battery mode.
- Power 230V(AC)/50Hz \pm 10%
- Stabilizer Inbuilt
- Digital display Digital display of time elapsed, set time,actual and set Pressure.
- Cuffs Different sizes of 5 cuffs-washable & easyfitting (pediatric, small, Medium, big & large.
- Flush Instantaneous Inflate & Deflate

Body Plastic

Alarm Activation along with display Indication for the following:

- After set time is over, it will show t.Ou.
- Only once it will show P.Hi, when pressure goes 25mm aboveset pressure.
- Only once it will show P.Lo, when pressure goes 25mmbelow set pressure.
- In battery mode, when battery voltage reaches 9V, it willshow bat. Lo.

Note:-

1. The bidder must quote **single Make**.
2. **Authorization certificate should be submitted.**
3. The bidder must quote their quotation only in the prescribed format (Annexure 2) on the letter head of firm otherwise quotation will be **REJECTED.**
4. Catalog must be attached with quotation for technical evaluation.
5. **The supplier may be asked for demonstration of their equipment** for which rates have been quoted, to the AIIMS Jodhpur. The expenditure incurred for demonstrating the items will be borne by the supplier.

On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

The Deputy Director (Admin.),
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR TOURNIQUET MACHINE AT AIIMS AGAINST THE INQUIRY NO. ADMIN/GEN/52-03/2022-AIIMS.JDH" DUE ON 27.06.2022 03.00 PM for Tourniquet Machine at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1	Tourniquet Machine	01						

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____