



Date: -08<sup>th</sup> June, 2017

Corrigendum  
For  
Tender for  
Hip Arthroscopy for the Department of Orthopaedics

NIT Issue Date	: 03 <sup>rd</sup> April, 2017
NIT No.	: Admn/Tender/04/2017-AIIMS.JDH
Pre-Bid Meeting	: 13 <sup>th</sup> April, 2017 at 03:45 PM
Earlier Last Date of Submission	: 15 <sup>th</sup> June, 2017 at 03:00 PM
Extended Last Date of Submission	: 21 <sup>st</sup> June, 2017 at 03:00 PM
Bid opening	: 22 <sup>nd</sup> June, 2017 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, In Technical Specification, Point No 1:**  
**For**  
Direct view, wide angle, Autoclavable, Hip Arthroscopes  
**Read**  
Direct view, wide angle, HD, Autoclavable, Hip Arthroscopes
- 2. Page No. 11, In Technical Specification, Point No 1, Direct view, wide angle, Autoclavable, Hip Arthroscopes, Sub Para No. 02:**  
**For**  
3.5-4.0mm – 70 degree, length 18-20 cm  
**Read**  
3.5-4.0mm – 70 degree, length 16-20 cm
- 3. Page No. 11, In Technical Specification, Point No 2, Instruments, Sub Para No. 02:**  
**For**  
Metallic cannula, diameter 4,5,6 mm  
**Read**  
Metallic cannula, diameter 4-4.8, 5-5.8, 6-6.8 mm