



Date: -14th July, 2017

Corrigendum
For
Tender for
Lithoclast for the Department of Pediatric Surgery

NIT Issue Date	: 11 th May, 2017
NIT No.	: Admn/Tender/101/2017-AIIMS.JDH
Pre-Bid Meeting	: 24 th May, 2017 at 03:00 PM
Earlier Last Date of Submission	: 08 th July, 2017 at 03:00 PM
Extended Last Date of Submission	: 02 nd August, 2017 at 03:00 PM
Bid opening	: 03 rd August, 2017 at 03:30 P.M

The following revised and additional specification will be added:-

1. Page No. 10, In Technical Specification, Point No. 2:

For

Combined - unit to generate energy for lithotripsy through shockwave during PCNL, cystoscopy and ureteroscopy; Adjustable pulse frequency; should work on power supply of 200 - 220 V; Ultrasound lithotripsy system; Ultrasound frequency 24 - 26 KHz

Read:

Combined - unit to generate energy for lithotripsy through shockwave during PCNL, cystoscopy and ureteroscopy; should work on power supply of 200 - 220 V; Ultrasound lithotripsy system; Ultrasound frequency 21 - 26 KHz; Mechanical energy 300 KHz

2. Page No. 10, In Technical Specification, Point No. 4:

For

Air compressor unit (integrated or separate) to generate pressure energy (8 bars) with tubing

Read:

Air compressor unit (integrated or separate) to generate pressure energy (8 bars) with tubing, as applicable

3. Page No. 10, In Technical Specification, Point No. 4:

For

Point No. 5 & 6: Handpiece for pneumatic lithotripsy with tubing - Two numbers; Handpiece for Ultrasound lithotripsy system with connection cable

Read:

The system should be supplied with two handpieces for pneumatic lithotripsy and one for ultrasound lithotripsy. In case both the energies are delivered through a common handpiece, two such handpieces must be supplied.

4. Page No. 10, In Technical Specification, Point No. 12:

For

Pneumatic probes (2 each) - a. 575 - 605 mm length; 1.6mm, 1mm & 0.8mm b. 375 - 500 mm length; 2mm, 1.6 - 3.2mm & 1mm

Read:

Pneumatic probes (2 each) - a. 575 - 605 mm length; 1.6mm, 1mm & 0.8mm b. 375 - 500 mm length; 2mm, 1.6 - 3.2mm & 1mm, where applicable

5. Page No. 10, In Technical Specification, Point No. 13:

For

Ultrasound probes (2 each) i. 400 - 403mm length; 3.5 - 3.8mm & 3 - 3.3mm ii. 300 - 330mm length; 3.3 - 3.5mm & 3.5 - 3.8mm iii. 570 - 573mm length; 1.5mm iv. 360 - 370 mm length; 1.5 - 1.9mm

Read:

Ultrasound probes (2 each) i. 396 - 403mm length; 3.5 - 3.8mm & 3 - 3.3mm ii. 300 - 396 length; 3.3 - 3.5mm & 3.5 - 3.8mm iii. 560 - 573mm length; 1.5mm iv. 360 - 418 mm length; 1.5 - 1.9mm

6. Page No. 10, In Technical Specification, Point No. 14:

For

All major items including trolley should be from same manufacturer

Read:

All major items including trolley should be from same manufacturer. If OEM is not manufacturing any part, a letter from the OEM should be enclosed with the bid and it could be provided from local manufacturer, but should be of similar quality.

7. Page No. 10, In Technical Specification, Point No. 18:

For

It should be US FDA/ European CE approved.

Read:

It should be US FDA and European CE approved. All the components including the main lithotripsy system should be from same manufacturer.