

Tender

For

Rigid Spinal HD Endoscopy System for the Department of
Neurosurgery

At

All India Institute of Medical Sciences, Jodhpur

NIT Issue Date	: 23 rd October, 2018
NIT No.	: Admn/Tender/104/2018-AIIMS.JDH
Pre-Bid Meeting	: 01 st November, 2018 at 03:15 PM
Last Date of Submission	: 27 th November, 2018 at 03:00 PM
Bid opening	: 28 th November, 2018 at 03:15 P.M

Tender documents may be downloaded from institute's web site www.aiimsjodhpur.edu.in (for reference only) and CPPP site <https://eprocure.gov.in/eprocure/app>



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur, Rajasthan-342005.

Telephone: 0291- 2740741, email: aoadmin@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

All India Institute of Medical Sciences (AIIMS), Jodhpur, Rajasthan, an apex healthcare institute established by an Act of Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites **Online bids in two bid system** for tenders for supply & installation of the Rigid Spinal HD Endoscopy System at the institute. You are requested to quote your best offer along with the complete details of specifications, terms & conditions.

Chapter-I

S.No	Item Description	Qty	EMD (Rs.)
1	Rigid Spinal HD Endoscopy System	01	Rs. 2,50,000/-

Instructions:

- Bids shall be submitted online only at CPPP website: <https://eprocure.gov.in/eprocure/app>.**
- The complete bidding process in online. Bidders should be possession of valid digital Signature Certificate (DSC) of class II or III for online submission of bids. Prior to bidding DSC need to be registered on the website mentioned above. For any assistance for e-bidding process, if required, bidder may contact to the helpdesk at 0291-2740741.
- Tenderer/Contractor/Bidders are advised to follow the instructions provided in the ‘Instructions to the Contractors/Tenderer/Bidders for the e-submission of the bids online through the Central Public Procurement Portal for e Procurement at <https://eprocure.gov.in/eprocure/app>’.**
- Bid documents may be scanned with 100 dpi with black and white option which helps in reducing size of the scanned document.

5. EMD Payment:

The bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount of **Rs. 2,50,000/- (Rupees Two Lakh Fifty Thousand Only)** by way of demand drafts or Bank Guarantee only. The demand drafts or Bank Guarantee shall be drawn in favour of “**All India Institute of Medical Sciences, Jodhpur**” payable at Jodhpur. The EMD of the successful bidder shall be returned after the successful submission of Bank Guarantee/ Security Deposit and for unsuccessful bidder(s) it would be returned after award of the contract. **The demand drafts or Bank Guarantee for EMD must deliver to AIIMS, Jodhpur on or before last date/time of Bid Submission.**

- Tenderer shall not be permitted to withdraw his offer or modify the terms and conditions thereof. In case the tenderer fail to observe and comply with stipulation made herein or backs out after quoting the rates, the aforesaid amount of earnest money will be forfeited.
- The Firm who are registered with National Small Industries Corporation (NSIC) / OR Small Scale Industries (SSI) are exempted to submit the EMD (Copy of registration must be provide along with technical bid)
- The EMD, in case of unsuccessful Bidders shall be retained by AIIMS, Jodhpur till the finalization of the tender. No interest will be payable by AIIMS, Jodhpur on the EMD.

6. Submission of Tender:

The tender shall be submitted online in two part, viz., technical bid and financial bid. All the pages of bid being submitted must be signed and sequentially numbered by the bidder irrespective of nature of content of the documents before uploading.

The offers submitted by Telegram/Fax/email shall not be considered. No correspondence will be entertained in this matter.

i) Technical Bid

The following documents are to be furnished by the Contractor along with **Technical Bid** as per the tender document:

- Signed and scanned copy of appropriate value of valid registration certificate (if any), experience certificate as per the tender notice, PAN, GST registration certificate and Tender Acceptance Letter.

- ii) Signed and Scanned Copy of Make and model of all systems, sub systems and additional items should be mentioned in the technical bid and complete technical details should be provided in the form of Brochures and write-ups.

Terms & Conditions:

1. **Validity:** The quoted rates must be valid for a period for 180 days from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.
2. **"PRE –BID Meeting" with the intending bidders shall be held on 01st November, 2018 from 03:15 PM. onwards at AIIMS, Jodhpur.** All the prospective bidders are requested to send comments/ representations on or before pre-bid meeting. Intending bidder will be allowed to seek clarification on specification, Conditions of Contract, etc. in writing to AIIMS, Jodhpur, within 48 hours after the pre-bid meeting.
3. In case the tenderer with Draws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.

4. Purchase Preference to Local Suppliers

In pursuance of Government of India Order No. P-45021/2/2017-B.E.-II dated 15th June 2017 and F.No. Z.28018/67/2017-EPW dated 12th June 2018 purchase preference shall be given to local suppliers in all procurements undertaken in the manner specified hereunder and the procurement shall be made as per terms and conditions contained in the said order.

5. **Minimum local content:** The minimum local content shall ordinarily be 25% as per Government of India Order No. P-45021/2/2017-B.E.-II dated 15/06/2017 and F.No. Z.28018/67/2017-EPW dated 12/06/2018, till the Nodal Ministry prescribes a higher or lower percentage. (Annexure-III).

6. **Margin of Purchase Preference:** The margin of purchase preference shall be 20%. The Local supplier whose quoted price falls in the margin of purchase preference desirous of claiming benefit of the Order No. P-45021/2/2017-B.E.-II dated 15/06/2017 shall submit an undertaking within 7 days of opening of financial bid, that he would be ready to supply the product at L1 price. In case of non-receipt of the same, he would not be given purchase preference.

7. The bidders are required to submit the following annexure in compliance of public procument (Preference to Make in India) order, 2017:

i) Affidavit of self-certification regarding local content (to be provided on Rs. 100/- stamp paper).

8. **All other terms & conditions will be as per the Department of Industrial Policy and Promotion (DTPP) order No. P-45021/2/2017-B.E.-II dated 15th June 2017.**

9. Delivery and Installation:

i) For goods supplied from India:

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **60 days** from the date of issue of supply order.

ii) For goods imported directly from abroad:

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **90 days** from the date of opening of Letter of Credit for shipment.

All the aspects of safe delivery, installation and commissioning shall be the exclusive responsibility of the supplier. If the supplier fails to deliver, install and commission the goods on or before the stipulated date, then a penalty at the rate of 0.5% per week of the total order value shall be levied subject to maximum of 10% of the total order value. The successful tenderer will also provide required training for supplied items at AIIMS Jodhpur. The goods should be manufactured after adoption of latest technology.

If at any time during the currency of the contract, the supplier encounters conditions hindering timely of the goods and performance of services, the supplier shall promptly inform the AIIMS, Jodhpur for extension of the delivery schedule accordingly. On receiving the supplier's communication, the AIIMS, Jodhpur shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.

In the case of package supply where the delayed portion of supply materially hampers installation and commissioning of the systems, liquidated damages charges shall be levied as above on the total value of the concerned package of the purchase order. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier.

- 8. Signing the Contract:** - The successful bidder shall be required to execute the Contract Agreement accepting all terms and conditions stipulated herein on a non-judicial stamp paper of Rs. 500/- (Rs. Five Hundred only) along with performance security within fifteen days of the issue of the Letter of notification of award. In the event of failure on the part of the successful bidder to sign the Contract within the period stipulated above, the EMD shall be forfeited and the acceptance of BID shall be considered as cancelled.
- 9. Performance Security:** As a guarantee towards due performance and compliance of the contract work, the successful bidder (contractor) will deposit an amount equal to 10% of order value and should be kept valid for a period of 60 day beyond completion of all the contractual obligation, including CMC period towards security deposit by way of demand draft/ bank Guarantee in favour of "All India Institute of Medical Sciences, Jodhpur" payable at Jodhpur drawn on any Nationalized Bank/Scheduled Bank and payable at Jodhpur within fifteen days of the issue of the Letter of notification of award along with non-judicial stamp paper of Rs. 500/- (Contract agreement).
- 10. Incidental Services:** The supplier shall be required to perform the following services:-
 - a. Installation & Commissioning, Supervision and Demonstration of the goods.
 - b. Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
 - c. On Site Training to Doctors/ Technicians/ Staff is to be provided by Supplier for operation and maintenance of the equipment for a period of 30 working days after successful installation of the machine, as per direction of user department.
 - d. Supplying required number of operation & maintenance manual for the goods.
 - e. To provide non-locked open software and standard interface inter-operability conditions for networked equipment's in hospital management information system, wherever applicable.
- 11. Accessories & Consumables:** The separate price list of all accessories and consumables, if any, must be attached/ enclosed along with the Financial Bid.
- 12. After Sales Service:** After sales service centre should be available on 24 (hrs.) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 24 hrs to ensure an uptime of minimum 95%, wherever applicable, failing which the necessary penalty measures shall be enforced.
- 13. Inspection:**
 - a. AIIMS, Jodhpur shall have the right to inspect and/or to test the goods to confirm their conformity to the NIT Specifications at no extra cost to the Purchaser.

- b. AIIMS, Jodhpur right to inspect, test and, where necessary, reject the Goods after the goods arrival at the final destination shall in no way be limited or waived by reason of the Goods having previously been inspected, tested and passed by AIIMS, Jodhpur prior to the goods shipment.
- c. The Director, AIIMS Jodhpur shall be the final authority to reject full or any part of the supply which is not confirming to the specification and other terms and conditions.
- d. No payment shall be made for rejected Stores. Rejected items must be removed by the Bidders within two weeks of the date of rejection at their own cost and replaced immediately. In case these are not removed, these will be auctioned at the risk and responsibility of the suppliers without any further notice.

14. Documents:

- a. **All pages of the Tender should be numbered and indexed.**
- b. The bidder shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully confirm to the goods and services specified by the purchaser in the tender documents. For this purpose the bidder shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the tender documents to establish technical responsiveness of the goods and services offered in its tender duly indicating relevant page numbers in the product literature.
- c. The bidder shall provide a list of major Government and Private Institutions where its relevant bid item has been supplied during last one year.

15. Manufacturer Authorisation: The bidder (if not original equipment manufacturer must submit Original Equipment Manufacturer authorization certificate that the tenderer is authorized for selling and maintain the equipment quoted for. Performa attached at **Annexure- IV**.

16. The bidders are required to submit user certificate for the relevant equipment on the letter head of the institution (Government/ Private).

17. The successful bidder will be required to submit order copies of the supply of the equipment in Government institutions in last 12 month for rate reasonability purpose.

18. Insurance: - The supplier shall make arrangements for insuring the goods against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery. If the equipment's is not commissioned and handed over to AIIMS, Jodhpur within specified period, the insurance will have to be extended by the supplier at their cost till the successful installation, testing, commissioning and handing over of the goods to the AIIMS, Jodhpur.

19. Tender Currencies:

- a. The bidder supplying indigenous goods or already imported goods shall quote only in Indian Rupees. Further, imported goods to be imported and supplied by the bidder are also required to be quoted in Indian Rupees.
- b. For imported goods if supplied directly from abroad, prices shall be quoted in any freely convertible currency say US Dollar, Euro, GBP or Yen. As regards price(s) for allied services, if any, required with the goods, the same shall be quoted in Indian Rupees only, if such services are to be performed /undertaken in India.
- c. Tenders, where prices are quoted in any other way shall be treated as non -responsive and rejected.

20. Tender Prices: While filling up the columns of the Financial Bid, the following aspects should be noted for compliance:

For domestic goods or goods of foreign origin located within India, the prices in the corresponding Financial Bid shall be entered separately in the following manner:

- a. The price of the goods, quoted ex-factory/ ex-showroom/ ex-warehouse/ off-the-shelf, as applicable, including all taxes and duties like GST, Custom Duty etc. already paid or payable on the components and raw material used in the manufacture or assembly of the goods quoted ex-factory etc. or on the previously imported goods of foreign origin quoted ex-showroom etc.;
- b. Any GST or other taxes, which will be payable on the goods in India if the contract is awarded;
- c. Charges towards Packing & Forwarding, Inland Transportation, Insurance, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Financial Bid;
- d. The price of Incidental Services, as mentioned in List of Requirements and Financial Bid;
- e. The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
- f. The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

For goods offered from abroad, the prices in the corresponding Financial Bid shall be entered separately in the following manner:

- a. The price of goods quoted FOB port of shipment, as indicated in the List of Requirements and Financial Bid;
- b. The price of goods quoted CIF port of entry in India as indicated in the List of Requirements and Financial Bid;
- c. The price of goods quoted for delivery at AIIMS, Jodhpur as indicated in the List of Requirements, Financial Bid and Consignee List;
- d. Wherever applicable, the amount of custom duty with CDEC applicable on CIF value on the goods to be imported;
- e. The charges for Loading/Unloading, Inland transportation, Insurance and other local costs, Incidental cost to delivery of the goods from the port of entry in India to AIIMS, Jodhpur, as specified in the List of Requirements and Financial Bid;
- f. The charges for Incidental Services, as in the List of Requirements and Financial Bid;
- g. The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
- h. The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

Additional information and instruction on Duties and Taxes: If the Bidder desires to ask for GST, Customs Duty etc. to be paid extra, the same must be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such GST and Customs duty etc. and no claim for the same will be entertained later.

GST:

- a. If reimbursement of GST is intended as extra over the quoted prices, the supplier must specifically state the same indicating the rate, quantum and nature of the GST applicable. In the absence of any such stipulation it will be presumed that the prices quoted are firm and final and no claim on account of GST will be entertained after the opening of tenders.
- b. If a bidder chooses to quote a price inclusive of GST and also desires to be reimbursed for variation, if any, in the GST during the time of supply, the Bidder must clearly mention the same and also indicate the rate and quantum of GST included in its price. No claim on account of GST will be entertained after the opening of tenders.
- c. Subject to sub clauses (i) & (ii) above, any change in GST upward/downward as a result of any statutory variation in GST taking place within contract terms shall be allowed to the extent of actual quantum of GST paid by the supplier. In case of downward revision in GST, the actual quantum of reduction of GST shall be reimbursed to the purchaser by the supplier. All such adjustments shall include all reliefs, exemptions, rebates, concession etc. if any obtained by the supplier.

Customs Duty: In respect of imported goods offered from abroad, the bidder shall specify the rate as well as the total amount of customs duty payable with Custom Duty Exemption Certificate, if applicable, on the quoted goods in the Financial Bid. The bidder shall also indicate the corresponding Indian Customs Tariff Number applicable for the goods.

- a. For transportation of imported goods offered from abroad, relevant instructions as incorporated shall be followed.
- b. For insurance of goods to be supplied, relevant instructions as provided shall be followed.
- c. Unless otherwise specifically indicated in this NIT document, the terms FCA, FOB, FAS, CIF, CIP etc. for imported goods offered from abroad, shall be governed by the rules & regulations prescribed in the current edition of INCOTERMS, published by the International Chamber of Commerce, Paris.
- d. The need for indication of all such price components by the bidders, as required in this clause is for the purpose of comparison of the tenders by the purchaser and will no way restrict the AIIMS, Jodhpur right to award the contract on the selected bidder on any of the terms offered.

21. Indian Agent:- If a foreign bidder has engaged an agent in India in connection with its bid, the foreign bidder, in addition to indicating Indian agent's commission, if any, shall also furnish the following information:

- a. The complete name and address of the Indian Agent and its Permanent Account Number as allotted by the Indian Income Tax authority.
- b. The details of the services to be rendered by the agent for the subject requirement.
- c. Details of Service outlets in India, nearest to the AIIMS, Jodhpur to render services during Warranty and CMC period.

22. Firm Price

- a. Unless otherwise specified in the NIT, prices quoted by the bidder shall remain firm and fixed during the currency of the contract and not subject to variation on any account.
- b. However, as regards taxes and duties, if any, chargeable on the goods and payable, the conditions stipulated will apply.

23. Conversion of tender currencies to Indian Rupees: - In case the bid document permits the bidders to quote their prices in different currencies, all such quoted prices of the responsive bidders will be converted to a single currency viz., Indian Rupees for the purpose of equitable comparison and evaluation, as per the closing exchange rates established by the Reserve Bank of India for similar transactions, as on the date of '**Last Date of Submission of Tender**'.

24. Payment Terms:

i) Payment for goods supplied from India:

100% payment of the total order value shall be released after the successful installation/commissioning of the ordered goods against the submission of the inspection report.

ii) Payment for Imported goods:

For imported goods payment shall be made in the following manner:

- a) **On shipment:** 75 % payment of the contract price shall be paid 60 days after presentation of shipping documents {goods shipped shall be paid through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the supplier in a bank in his country} and upon the submission of the following documents:
 - i. Four copies of Supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
 - ii. Original and four copies of the clean, on-board Bill of Lading/ Airway bill, marked freight prepaid and four copies of non-negotiable Bill of Lading/Airway bill.
 - iii. Insurance Certificate;
 - iv. Certificate of origin by the chamber of commerce of the concerned country;
 - v. Certificate of country of origin;
 - vi. Manufacturer's / Supplier's warranty certificate;
 - vii. Manufacturer's own factory inspection report.

- b) **On Acceptance:** 25 % payment would be made after satisfactory installation & commissioning on issuance of Inspection certificate by the AIIMS, Jodhpur.

Note:-The supplier shall not claim any interest or any other payment under the contract.

- 25. Custom Clearance:** For the Goods to be imported and supplied, the Institute will provide Custom Duty Exemption Certificate (CDEC) to successful bidder for availing concessional rate of duty as per prevailing Custom Tariff. In case, the bidder requires CDEC certificate, then the same should be specifically mentioned in the bid. The supplier is solely responsible for getting the material clearance from customs. Institute will provide all custom documents for custom clearance on the demand of supplier. Transportation of goods up to AIIMS, Jodhpur and its successful installation and commissioning is also the responsibility of the supplier. All charges/ expenses incurred in this process will be borne by the supplier. **NO DEMURRAGE / WHARFAGE CHARGES WILL BE PAYABLE BY THE INSTITUTE UNDER ANY CIRCUMSTANCES. NO ADVANCE PAYMENT WILL BE PAYABLE FOR CUSTOM CLEARANCE/ FREIGHT/INSURANCE ETC.**
- 26. Guarantee / Warrantee Period:** The Tenderers must quote for **05 years** comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **05 years** Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.
- 27. Uptime guarantee:**The firm should provide uptime guarantee of 95%
- 28. Downtime penalty Clause**
- a. During the comprehensive warranty period, the guarantee uptime of 95% of 365 days will be ensured. In case the down time exceeds the 5% limit penalty of extension of guaranty period by two days for each additional day of down time will be enforced. The vendor must undertake to supply all spares for optimal upkeep of the equipment for at least FIVE YEARS after handling over the unit to the Institute. If accessories / other attachment of the system are procured from the third party, then the vendor must produce cost of accessory / other attachment and the CMC from the third party separately along with the main offer and the third party will have to sign the CMC with the Institute if required.
- b. The principals or their authorized service providers are required to submit a certificate that they have satisfactory service arrangements and fully trained staff available to support the uptime guarantee.
- 29. Arbitration:** If any difference arises concerning this agreement, its interpretation on payment to the made there-under, the same shall be settled out by mutual consultation and negotiation. If attempts for conciliation do not yield any result within a period of 30 days, either of the parties may make a request to the other party for submission of the dispute for decision by an Arbitral Tribunal containing Sole Arbitrator to be appointed by the Secretary, Department of Legal Affairs. Such requests shall be accompanied with a panel of names of three persons to act as the sole arbitrator. In case of such arbitrator refusing, unwilling or becoming incapable to act or his mandate having been terminated under law, another arbitrator shall be appointed in the same manner from among the panel of three persons to be submitted by the claimant. The provision of Arbitration and Conciliation Act, 1996 and the rule framed there under and in force shall be applicable to such proceedings.
- 30. Subletting of Work:** The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Jodhpur, which will be at liberty to refuse if thinks fit. The tender is not transferable. Only one tender shall be submitted by one tenderer.

- 31. Breach of Terms and Conditions:** In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order/ job without assigning any reason thereof and nothing will be payable by AIIMS, Jodhpur in that event the security deposit shall also stands forfeited.
- 32. Insolvency etc:** In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other order under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Jodhpur shall have the power to terminate the contract without any prior notice.
- 33. Force Majeure:** If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, explosion, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party shall be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries have been so resumed or not shall be final and conclusive.
- Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, either party may, at least option to terminate the contract.
- 34.** Bidder shall submit a copy of the tender document and addenda thereto, if any, with each page of this document should be signed and stamped to confirm the acceptance of the entire terms & conditions as mentioned in the tender enquiry document.
- 35.** The quantity of item given in the tender is tentative, which may be increased or decreased as per the institute's requirement.
- 36.** Signed & stamped compliance sheet of the technical specification of the goods with technical printed literature must be enclosed with the bid.
- 37.** After due evaluation of the bid(s) Institute will award the contract to the lowest evaluated responsive tenderer
- 38.** Conditional bid will be treated as unresponsive and it may be rejected.
- 39. Demonstration:** - AIIMS Jodhpur reserves the right to ask the tenderers for arranging demonstration of their equipment for which rates have been quoted, to the concerned committee, if required.
- 40.** The Institute reserves the right to accept in part or in full or reject any or more tender(s) without assigning any reasons or cancel the tendering process and reject all tender(s) at any time prior to award of contract, without incurring any liability, whatsoever to the affected bidder or bidder(s).
- 41. Applicable Law:**
- The contract shall be governed by the laws and procedures established by Govt. of India, within the framework of applicable legislation and enactment made from time to time concerning such Commercial dealings / processing.
 - Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Jodhpur, Rajasthan, India only.
 - The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jodhpur. The decision of the Arbitrator shall be final and binding on both the parties.
 - Force Majeure: Any delay due to Force Majeure will not be attributable to the supplier.

Administrative Officer
AIIMS, Jodhpur

Annexure-I

S.No	Item	Technical Specification	Qty																						
01.	Rigid Spinal HD Endoscopy System	<p>Technical Specifications “Rigid spinal HD Endoscopy System” The system is required for endoscopic spine surgeries which includes endoscopic lumbar and lumbosacral discectomies, endoscopic spinal canal decompressions and endoscope assisted minimal invasive spine surgeries.</p> <p>A. Endoscopic Spine Set for Disc Herniation B. Endoscopic Spine Set with sequential dilators and scope C. Video Neuro Endoscope (flexible telescope with integrated camera head) D. Exoscope with Pneumatic Universal holding arm</p> <p>A – Endoscopic Spine Set for Disc Herniation</p> <table border="1" data-bbox="464 636 1353 2125"> <thead> <tr> <th data-bbox="464 636 1233 824">Spine Set for central to far lateral disc herniation and lumbar spinal stenosis. Set must be mobile device to give freedom of cranial, cordial, medial & lateral movement. Endoscopy Set must be US FDA and European CE approved.</th> <th data-bbox="1233 636 1353 824"></th> </tr> <tr> <th colspan="2" data-bbox="464 824 1353 862">Technical Specifications:</th> </tr> <tr> <th data-bbox="464 862 1233 900">Item</th> <th data-bbox="1233 862 1353 900">Qty</th> </tr> </thead> <tbody> <tr> <td data-bbox="464 900 1233 1111">Localization Device for Lumbar Spinal Cord. A metal unique device for localization of spinal vertebra in fluoroscopy for determination of the point of incision. It should be able to cover from pediatric to adult patient It should be able to cover from lean to very over weight patient</td> <td data-bbox="1233 900 1353 1111">1</td> </tr> <tr> <td data-bbox="464 1111 1233 1279">Straight Forward Telescope 0°, Should be with enlarged view, Should have diameter 4 mm, Should have length 18 cm, Should be autoclavable, Should have fiber optic light transmission incorporated</td> <td data-bbox="1233 1111 1353 1279">2</td> </tr> <tr> <td data-bbox="464 1279 1233 1417">Operating Tube for posterior should have inner diameter line spinal endoscopy , tube should be oval in shape should supplied with obturator</td> <td data-bbox="1233 1279 1353 1417">2</td> </tr> <tr> <td data-bbox="464 1417 1233 1686">Working tube should come with Height adjustment option should come with inbuilt detachable adjustable Nerve Protector/retractable, should be used with operating tube should have 3 channels: 1. scope fixing channel 2. operating channel 3. separate channel for suction tube should have minimum working channel of 8mm.</td> <td data-bbox="1233 1417 1353 1686">2</td> </tr> <tr> <td data-bbox="464 1686 1233 1789">Chisel should be flat, straight, with non-slippery handle, should have tip from 14 to 15 mm, should have working length 8.5 to 9 cm</td> <td data-bbox="1233 1686 1353 1789">1</td> </tr> <tr> <td data-bbox="464 1789 1233 1892">Suction Tube should be angled, with suction control hole should have diameter 3.6 to 3.7 mm, should have working length 11 to 11.5 cm</td> <td data-bbox="1233 1789 1353 1892">2</td> </tr> <tr> <td data-bbox="464 1892 1233 2060">Spoon/disc forceps should be dismantling should have oval in shape with only single jaws moving. Should have spoon size 3 x 10 mm should have working length 14.5 to 15 cm</td> <td data-bbox="1233 1892 1353 2060">2</td> </tr> <tr> <td data-bbox="464 2060 1233 2125">Elevator, should have spatula shape with slightly curved tip , should have distal tip should have inner diameter 4.7 to 5 mm,</td> <td data-bbox="1233 2060 1353 2125">1</td> </tr> </tbody> </table>	Spine Set for central to far lateral disc herniation and lumbar spinal stenosis. Set must be mobile device to give freedom of cranial, cordial, medial & lateral movement. Endoscopy Set must be US FDA and European CE approved.		Technical Specifications:		Item	Qty	Localization Device for Lumbar Spinal Cord. A metal unique device for localization of spinal vertebra in fluoroscopy for determination of the point of incision. It should be able to cover from pediatric to adult patient It should be able to cover from lean to very over weight patient	1	Straight Forward Telescope 0°, Should be with enlarged view, Should have diameter 4 mm, Should have length 18 cm, Should be autoclavable, Should have fiber optic light transmission incorporated	2	Operating Tube for posterior should have inner diameter line spinal endoscopy , tube should be oval in shape should supplied with obturator	2	Working tube should come with Height adjustment option should come with inbuilt detachable adjustable Nerve Protector/retractable, should be used with operating tube should have 3 channels: 1. scope fixing channel 2. operating channel 3. separate channel for suction tube should have minimum working channel of 8mm.	2	Chisel should be flat, straight, with non-slippery handle, should have tip from 14 to 15 mm, should have working length 8.5 to 9 cm	1	Suction Tube should be angled, with suction control hole should have diameter 3.6 to 3.7 mm, should have working length 11 to 11.5 cm	2	Spoon/disc forceps should be dismantling should have oval in shape with only single jaws moving. Should have spoon size 3 x 10 mm should have working length 14.5 to 15 cm	2	Elevator, should have spatula shape with slightly curved tip , should have distal tip should have inner diameter 4.7 to 5 mm,	1	01
Spine Set for central to far lateral disc herniation and lumbar spinal stenosis. Set must be mobile device to give freedom of cranial, cordial, medial & lateral movement. Endoscopy Set must be US FDA and European CE approved.																									
Technical Specifications:																									
Item	Qty																								
Localization Device for Lumbar Spinal Cord. A metal unique device for localization of spinal vertebra in fluoroscopy for determination of the point of incision. It should be able to cover from pediatric to adult patient It should be able to cover from lean to very over weight patient	1																								
Straight Forward Telescope 0°, Should be with enlarged view, Should have diameter 4 mm, Should have length 18 cm, Should be autoclavable, Should have fiber optic light transmission incorporated	2																								
Operating Tube for posterior should have inner diameter line spinal endoscopy , tube should be oval in shape should supplied with obturator	2																								
Working tube should come with Height adjustment option should come with inbuilt detachable adjustable Nerve Protector/retractable, should be used with operating tube should have 3 channels: 1. scope fixing channel 2. operating channel 3. separate channel for suction tube should have minimum working channel of 8mm.	2																								
Chisel should be flat, straight, with non-slippery handle, should have tip from 14 to 15 mm, should have working length 8.5 to 9 cm	1																								
Suction Tube should be angled, with suction control hole should have diameter 3.6 to 3.7 mm, should have working length 11 to 11.5 cm	2																								
Spoon/disc forceps should be dismantling should have oval in shape with only single jaws moving. Should have spoon size 3 x 10 mm should have working length 14.5 to 15 cm	2																								
Elevator, should have spatula shape with slightly curved tip , should have distal tip should have inner diameter 4.7 to 5 mm,	1																								

	should have working length 13 cm		
	Hook should be blunt should be distally angled 90°, should have hook length 5.5 mm to 5.4mm, should have working length 13 to 14 cm	1	
	Trephine for puncturing disc and for shaving bone should be with round handle should have diameter 2.8 to 3 mm, should have working length 22 to 22.5 cm	2	
	Bone Punch, should have dismantling option, should have 90° up-biting cut, should have bite size 3 mm, should have working 17.5 to 18.5cm	2	
	Bone Punch, should have dismantling option , should have up-biting 45° forward, should have bite size should have 3 mm, should have working 17.5 to 18.5cm	2	
	Bipolar Forceps, should have distill tip rounded , should have tip should have inner diameter 1.9 to 2 mm, should have outer diameter 3.3 to 3.4 mm, should have working length 19 to 20 cm.	2	
	Bipolar Forceps, Should have rounded tip, Should have tip should have inner diameter 2 mm, Should have outer diameter 3.4 mm, Should have working length 19 to 20 cm	1	
	Bipolar High Frequency Cord with 2 x 4 mm banana-plug to Coagulator , length minimum 300 cm	2	
	Ball Probe, should have 4 mm diameter ball at the distill end of probe should be straight with working length 13 to 13.5 cm	1	
	Hook for palpation of nerve , should be blunt with distal end double curved should have working length 13 to 13.5 cm	1	
	Localization Device for Cervical Spinal Cord a metal unique device for localization of spinal vertebra in fluoroscopic light for determination of the point of incision this should be able to cover from pediatric to adult patient this should be able to cover from lean to very over weight patient	1	
	Tissue Retractor Blade with length 16 to 17 cm, should be double-ended, should have size 35 mm x 10 mm and 21 mm x 10 mm of both end	4	
	Retractor Blade for tissue should be conical in shape should have length 4.5 to 5 cm	4	
	Bone Punch should be detachable in 2 pieces for cleaning, should have up-biting 90° forward cut with bite size 2 mm, should have working length 16.5 to 17 cm	2	
	Bone Punch should be detachable in 2 pieces for cleaning should have up-biting 40° forward cut with bite size 2 mm, should have working length 16.5 to 17 cm	2	

		<p>Plastic Container for Sterilization and Storage of Variable Instrument Sets, Should be perforated, with transparent should have inner diameter, Should be with silicone mat, should be having two-level storage Should have minimal external dimensions (w x d x h): 545 x 260 x 115 mm The plastic containers Should be suitable for steam, gas and hydrogen peroxide</p>	2		
		<p>B. Endoscopic Spine Set with sequential dilators and scope</p>			
		<p>Surgical Write up - Endoscopic Discectomy, Decompression of the Spinal Canal. Universal system for minimally invasive lumbar and dorsal cervical disc and spinal surgery. System must combine all the advantages of microsurgery and spinal endoscopy. The use of a dilator system in conjunction with various size trocars must affords the main advantage of minimal incisions and least muscle trauma. Endoscopy Set must be US FDA and European CE</p>			
		<p>Technical Specifications:</p>			
		<p>Dilation Sleeve Set of 7 dilators should be graduated, 1. Inner diameter 1.5 mm, Outer diameter 5.2 mm, length 23 cm. 2. Should have outer diameter 8.9 mm, should have inner diameter 5.3 mm, length 21 cm. 3. Inner diameter 9 mm, outer diameter 12.7 mm, length 19 cm. 4. Should have outer diameter 14.9 mm 5. Should have inner diameter 12.9 mm, graduated, length 17 cm. 6. Should have outer diameter 16.9 mm, should have inner diameter 15.1 mm, graduated, length 15 cm. 7. Should have outer diameter 18.9 mm, should have inner diameter 17.1 mm, graduated, length 14 cm. 8. Should have outer diameter 20.9 mm, should have inner diameter 19 mm, graduated, length 13 cm. Telescope 25° with eyepiece angled at 90° with no working channel, telescope should have diameter 4 mm, telescope should come in length 6 cm should be autoclavable with fiber optic light transmission incorporated telescope should be used with 4 -4.5 cm length trocar set</p>	Set of 7 (1 dilator each)		
		<p>Telescope 25° with eyepiece angled at 90° with no working channel, telescope should have diameter 4 mm, telescope should come in length 6 cm should be autoclavable with fiber optic light transmission incorporated telescope should be used with 7 -7.5 cm length trocar set</p>	1		
		<p>Telescope 25° with eyepiece angled at 90° with no working channel, telescope should have diameter 4 mm, telescope should come in length 6 cm should be autoclavable with fiber optic light transmission incorporated telescope should be used with 9 -9.5 cm length trocar set</p>	2		
		<p>Telescope 25° with eyepiece angled at 90° with no working channel, telescope should have diameter 4 mm, telescope should come in length 6 cm should be autoclavable with fiber optic light transmission incorporated telescope should be used with 9 -9.5 cm length trocar set</p>	1		

	<p>Trocar Set should come in 9 different combinations</p> <ol style="list-style-type: none"> 1. trocar diameter 15 mm Trocar length 40mm 2. trocar diameter 15 mm Trocar length 70mm 3. trocar diameter 15 mm Trocar length 90 mm 4. trocar diameter 19mm Trocar length 40 mm 5. trocar diameter 19 mm Trocar length 70 mm 6. trocar diameter 19 mm Trocar length 90 mm 7. trocar diameter 23 mm Trocar length 45 mm 8. trocar diameter 23 mm Trocar length 75mm 9 trocar diameter 23 mm Trocar length 95 mm 	Set of 9 (1 trocar each)		
	Accessory/attachment/scope holders for 15, 19 and 23 mm trocar set of 6	Set of 6 (1 each)		
	Bone Punch should be detachable for cleaning 40° forward up biting with bite size 4 mm, working length 23.5 to 24 cm	1		
	Bone Punch should be detachable for cleaning should be , 40° up biting, with bite size 2 mm, working length 23.5 to 24 cm	1		
	Bone Punch should be detachable for cleaning should be 90° up biting with bite size 2 mm, working length 23.5 to 24 cm	1		
	Bone Punch should be detachable for cleaning should be 90° up biting, with bite size 4 mm, working length 23.5 to 24 cm	1		
	Bone Punch should be detachable for cleaning should be bayonet shape down biting 40° forward, with bite size 2 mm, should have working length 16.5 to 17 cm	2		
	Suction Tube, should be with cut-off hole for suction control, working length 14.5 to 15 cm, diameter 2.5 to 2.7 mm	1		
	Suction Tube, should be with distal nerve retractor along with cut-off hole for suction control should have diameter 2.5 - 2.7 mm, working length 14.5 to 15 cm	1		
	Nerve Retractor, should be angled 30°, distal should have inner diameter 5 mm, should have working length 16.5 to 17 cm	1		
	Dissector with sharp, tip angled 15°, should have round handle, tip distill size should be 2 mm, should be having length 24-25 cm	1		
	Palpation Hook, should be with distally 10 mm long with angled 90°, with ball at the end of tip, working length 19 to 20 cm	1		
	Spoon Forceps with oval shape jaw. spoon size 3 x 10 mm, should have single action jaws, working length 19 to 20 cm	1		
	Scissors with hook shape jaw with single action jaws, size 2.7 to 2.7 mm, working length 25-26 cm	1		
	Socket to clamp on the operating table, Should be able to use with European and United States standard rails, Should also be suited for rails from 25x10 up to 35x8 mm, with lateral clamping element for height adjustment of the articulated stand should be rotatable	1		
	Articulated Stand, holding arm should be mechanical L shaped.	1		

	Should have one mechanical central clamp for all minimum five or more joints joint moving functions, Should have height 48 to 50 cm, Should have operating range 50 to 52 cm		
	Nerve Retractor with hook length 1.9-2 mm, diameter 3.9-4 mm, with angled sheath, working length 19 to 20 cm	1	
	Nerve Hook should have inner diameter 5 mm, should be bayonet in shape wit length 15.5-16 cm	1	
	Nerve Hook should have inner diameter 3 mm, should be bayonet in shape wit length 15.5-16 cm	1	
	Palpation Hook should be bayonet shaped, should be distally angled 90° with ball end, with round handle, working length 19 to 20 cm	1	
	Surgical handle, should be bayonet-shaped with working length 15-15.5 cm.	1	
	Nucleus Cutting Forceps should have single action jaws should have diameter 3.3-3.5 mm, working length 19 to 20 cm.	1	
	Bipolar Forceps, Should have rounded tip, Should have tip should have inner diameter 2 mm, Should have outer diameter 3.4 mm, Should have working length 19 to 20 cm	2	
	Bipolar High Frequency Cord with 2 x 4 mm banana-plug to Coagulator , length minimum 300 cm	1	
	Plastic Container for Sterilization and Storage of Variable Instrument Sets, Should be perforated, with transparent should have inner diameter, Should be with silicone mat, should be having two-level storage Should have minimal external dimensions (w x d x h): 545 x 260 x 115 mm The plastic containers Should be suitable for steam, gas and hydrogen peroxide	3	
	C. Video Neuro Endoscope (flexible telescope with integrated camera head)		
	Video neuro should have inner diameter Neuro Endoscope, should exhibit an exceptionally thin and flexible sheath also Should be able to able to transmit outstanding image quality when combined with HD camera system. Should be able to perform endoscopy of the subdural and subarachnoid space. Should have angle of view 270 degree. should have excelled image quality with light cables incorporated in scope Should feature digital zoom minimum up to 2x magnification to be able to improve detail resolution for high precision work. Should have integrated working channel to allow intraoperative use of flexible instrument size up to 1mm. Technical specifications: Should be flexible, Should be steerable, direction of view 0°, angle of view 90°, Should have working channel maximum size 3.5 fr to 3.6 Fr. (approx. 1.2mm) Should have sheath size maximum sheath size 8.4 Fr to 8.5 Fr., (approx. 2.8mm) Should have Working length 32 to 35 cm	1	

	Should be supplied with storage case. Should be GAS and plasma sterilisable. Flexible scope and camera head must be one single piece. This should directly get attached with camera control unit through camera cable.		
	Biopsy Forceps, Should have double action jaws, Should be flexible, Should have diameter maximum 1.0 mm, Should have length minimum 60 cm	1	
	Grasping Forceps, Should be double action jaws, Should be flexible with diameter maximum 1.0 mm and length minimum 60 cm	1	
	Coagulating Electrode should have outer diameter, unipolar, flexible, 1.0 mm, length 73 cm	1	
	Unipolar High Frequency Cord, with 4 mm plug, length minimum 300 cm, For use with HF units .	1	
	D. Exoscope with Pneumatic Universal holding arm		
	Technical / surgical Write up - Pneumatic Universal Medical Holding Arm along with, Exoscope a unique visualization system with for open surgery with minimal access – Telescope for visualization magnification and broadcasting of all neuro and spine surgical procedures		
	TECHNICAL SPECIFICATION:		
	An OT table mounted holding system to hold endoscopes retractors or instruments during surgery. The holding arm should be safe, easy to handle. Holding arm should be pure - pneumatic with no additional control unit one easy button handling at the distill end of arm must be easy to handle and with one button locking mechanism it should have OR table adapter (to be fixed with OR table railing) it should have 600 to 650 mm length for large working area It should be able to hold between 1.8 kg to 2 kg force. Technical Specifications for ARM No of Joints in Arm - Minimum 5 joints plus at least 1 manual ball joint Arm length - 600 to 650mm, weight of Arm - 5.5 to 6.5 kg, maximum load for arm - 1.8 to 2 KG Arm Operating type / mechanism - should be pneumatic with self-locking (no electrical support required) Operating Gas for Arm - N2 or compressed air. Gas pressure for Arm - min 6 to max 8 BAR (90 to 115 Psi). Sterilization procedure for ARM- should be able to drape in sterile cover OR table adapter weight - 6-7 KG OR table adapter joint - 3 to 5 Should be supplied with minimum 3-4 different size clamp with holding range from 0 to 23 mm also clamp should be able to hold fiberscopes easily. Should be supplied with air compressor and compressor connecting tubes with pressure regulator. Air Compressor should have power supply 230 VAC, 50/60 Hz Air compressor should have Dimensions maximum (w x d x h): ca. 500 x 320 x 340 mm and weight maximum 23 kg Air compressor should have - Air compressor & Fabric Reinforced tube for connecting, Connecting Tube, length minimum 600 cm	1 each	

		Required Pressure setting for air compressor should be min. 6 to max. 8 bar.			
		Exoscope with either 0 degree or 90 degree view Should have inbuilt light transmission rod lenses with 2 light outlet point at the tip of scope Should be used in working distance minimum 25 - maximum 75 cm, not to be used with close contact with patient Length should be 10 - 11 cm, should be autoclavable When scope apparatus kept at min height of 25 cm depth of field should be in the range of 3 to 4 cm When scope apparatus kept at max height of 75 cm depth of field should be in the range of 9.5 to 10.5 cm	1		
		Fiber Optic Light Cable should have with straight connector Should be extremely heat-resistant with enhanced light transmission Should have diameter 4.6 to 4.8 mm, Should have length 300 to 310 cm	2		
		Clamping Cylinder An autoclavable plastic folding cylinder which can be opened vertically to be fixed on 10 mm telescope. This should be able to be used as sheath to 10mm telescope with cylinder length not more than 4cm for 10-11 cm scope	2		
		Clamping Jaw, Should be of metal, Should be for use with instrument and telescope sheaths, clamping range 16.5 up to 23 mm.	1		

Annexure-II**TECHNICAL BID**

Name of Firm/Contractor/Supplier	
Complete Address & Telephone No.	
Name of Proprietor/Partner/Managing Director/Director.	
Phone No:- Mobile No:- Email Id:-	
Name and address of service centre nearby Jodhpur.	
Whether the firm is a registered firm Yes/No (attached copy of certificate).	
PAN No. (enclose the attested copy of PAN Card).	
GST IN (enclose the attested copy of GST Registration Certificate).	
Whether the firm has enclosed the Bank Draft/Pay Order/Banker's cheque of Earnest Money Deposit.	
Whether the Firm/Agency has signed each and every page of Tender/NIT.	
Please provide full list of consumables.	
Any other information, if necessary.	

Authorized signatory of the bidder with seal.

Annexure-III**Calculation of Local Content**

Name of manufacturer	Calculation by Manufacturer (Cost per unit of product)		
Cost Component	Cost (Domestic Component) a	Total Cost b	Percentage of Local Content $c=(a/b)*100$
I.			
II.			
III. Total Cost (Excluding tax and duties)			

1. **Cost (Domestic Component):** Cost of domestic component may be calculated based on one of the followings depending on data available. Each of these calculations should provide consistent result.
 - a. Sum of the costs of all inputs which go into the product (including duties and taxes levied on procurement of inputs except those for which credit/ set-off can be taken) and have not been imported directly or through a domestic trader or an intermediary.
 - b. Ex-Factory Price of product minus profit after tax minus sum of imported Bill of Material used (directly or indirectly) as inputs in producing the product (including duties and taxes levied on procurement of inputs except those for which: credit/ set-off can be taken) minus warranty costs.
 - c. Market price minus post-production freight, insurance and other handling costs minus profit after tax minus warranty costs minus sum of Imported Bill of Material used as inputs in producing the product (including duties and taxes levied on procurement of inputs except those for which credit / set-off can be taken) minus sales and marketing expenses.

2. **Total Cost:** Total cost may be calculated based on one of the following depending on data available. Each of these calculations should provide consistent result.
 - a. Sum of the costs of all inputs which go into the product (including duties and taxes levied on procurement of inputs except those for which credit / set-off can be taken).
 - b. Ex-Factory Price of product minus profit after tax, minus warranty costs.
 - c. Market price minus pent-production freight, insurance and other handling costs minus profit after tax, minus warranty costs minus sales and marketing expenses.

Format for Affidavit of Self Certification regarding Local Content in a Medical Devices to be purchase on Rs. 100/- Stamp Paper.

I _____ S/o, D/o, W/o _____ Resident
of _____ do

hereby solemnly affirm and declare as under:

That I will agree to abide by the terms and conditions of the policy of Government of India issued vide Notification No:

That the information furnished hereinafter is correct to best of my knowledge and belief and I undertake to produce relevant records before the procuring entity or any authority so nominated by the Department of Pharmaceuticals. Government of India for the purpose of assessing the local content.

That the local content for all inputs which constitute the said medical device has been verified by me and I am responsible for the correctness of the claims made therein.

That in the event of the domestic value addition of the product mentioned herein is found to be incorrect and not meeting the prescribed value-addition norms based on the assessment of an authority so nominated by the Department of Pharmaceutical. Government of India for the purpose of assessing the local content, action will be taken against me as per Oder No. P-45021/2/2017-B.E-II dated 15.06.2017 and Guidelines issued vide letter no. 31026/36/2016- MD dated – 18.05.2018.

I agree to maintain the following information in the company's record for a period of 8 years and shall make this available for verification to any statutory authority.

- i. Name and details of the Domestic Manufacturer (Registered Office, Manufacturing unit location, nature of legal entity).
- ii. Date on which this certificate is issued.
- iii. Medical devices for which the certificate is produced
- iv. Procuring entity to whom the certificate is furnished
- v. Percentage of local content claimed
- vi. Name and contact details of the unit of the manufacturer
- vii. Sale Price of the product
- viii. Ex-Factory Price of the product
- ix. Freight, insurance and handling
- x. Total Bill of Material
- xi. List and total cost value of inputs used for manufacture of the medical device.
- xii. List and total cost of inputs which are domestically sourced Value addition certificates from suppliers. If the input is not in use attached.
- xiii. List and cost of inputs which are imported, directly or indirectly.

For and on behalf of

(Name of firm/entity)
Authorized signatory

Annexure-IV

MANUFACTURER'S / PRINCIPAL'S AUTHORIZATION FORM

To

The Administrative Officer,
All India Institute of Medical Sciences, Jodhpur

Sir,

TENDER: _____.

we, _____, who
are established and reputable manufacturers of _____,
having factories at _____ and _____,
hereby authorize Messrs. _____ (name and address of
agents)

to bid, negotiate and conclude the contract with you against Tender

No. _____ for the above goods manufactured by
us. No company or firm or individual other than Messrs. _____ are
authorized to bid, negotiate and conclude the contract in regard to this business
against this specific tender.

We hereby extend our full guarantee and warranty as per the conditions of tender for
the goods offered for supply against this tender by the above firm.

The authorization is valid up to _____

Yours faithfully,

(Name)

For and on behalf of Messrs. _____
(Name of manufacturers)/Principal.

Annexure-V
Financial Bid

A) FINANCIAL BID FOR DOMESTIC GOODS OR GOODS OF FOREIGN ORIGIN LOCATED WITHIN INDIA OR GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES

1	2	3	4	6						
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	Ex - factory/ Ex-warehouse /Ex-showroom /Off-the shelf	Packing and Forwarding charges	Inland Transportation, Insurance, loading/unloading and Incidental costs at AIIMS-Jodhpur	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS-Jodhpur	GST RATE [%age & value]	Unit Price (at AIIMS-Jodhpur) basis	Total Price (at AIIMS-Jodhpur) basis (Rs.)
				(a)	(b)	(c)	(d)	(e)	(f)= (a+b+c+d+e)	= {4 x 5(f)}

Total Tender price in Rupees: _____

In words: _____

Note: -

1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
2. The charges for Annual CMC after warranty shall be quoted separately.
3. The Bidder must quote price for "GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES" after having taken in to account, the provision of Custom Duty Exemption Certificate (CDEC) by the Purchaser, as per Customs Tariff Act.

Place:

Date:

Name:

Business Address:

Signature of Bidder:

Seal of the Bidder:

Financial Bid**B) FINANCIAL BID FOR GOODS TO BE IMPORTED FROM ABROAD**

1	2	3	4	5				6
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	Price per unit				Total price on Destination + Insurance (local transportation and storage) = {4X 5 (d)}
				FOB price at port/ airport of Lading (a)	Carriage & Insurance (port of loading to port of entry) and other Incidental costs** (b)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS-Jodhpur ** (c)	Unit Price on DDP AIIMS-Jodhpur + Extended Insurance (local transportation and storage) (d) = a+b+c	

** To be paid in Indian Currency (Rs.)

Total Tender price in foreign currency: _____

In words: _____

Note: -

1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
2. The charges for Annual CMC after warranty shall be quoted.
3. The Bidder will be fully responsible for the safe arrival of the goods AIIMS-Jodhpur in good condition as per terms of DDP as per INCOTERMS, if applicable.

Indian Agent:

Indian Agency Commission - ____% of FOB

Place:

Date:

Name:

Business Address:

Signature of Bidder:

Seal of the Bidder:

Financial Bid

C) FINANCIAL BID FOR ANNUAL COMPREHENSIVE MAINTENANCE CONTRACT AFTER WARRANTY PERIOD:

1. S.No.	2. DESCRIPTION OF GOODS	3. QUANTITY. (Nos.)	4. Annual Comprehensive Maintenance Contract Cost for Each Unit year wise*.					5. Total Annual Comprehensive Maintenance Contract Cost for 5 Years [3 x (4a+4b+4c+4d+4e)]
			1 st	2 nd	3 rd	4 th	5 th	
			a	b	c	d	e	

* After completion of Warranty period.

GST: Whether extra or inclusive, if extra, indicates the rate_____.

NOTE:-

1. In case of discrepancy between unit price and total prices, THE UNIT PRICE shall prevail.
2. The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual, labour and spares, after satisfactory completion of Warranty period may be quoted as per NIT conditions on yearly basis for complete equipment and Turnkey (if any).
3. The cost of CMC may be quoted along with taxes applicable. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
4. Cost of CMC will be added for Ranking/Evaluation purpose.
5. All software updates should be provided free of cost during CMC period.
6. The stipulations in Technical Specification will supersede above provisions
7. The supplier shall keep sufficient stock of spares required during Annual Comprehensive Maintenance Contract period. In case the spares are required to be imported, it would be the responsibility of the supplier to import and get them custom cleared and pay all necessary duties.

Date:

Place:

Name:

Business Address:

Signature of Bidder:

Seal of the Bidder: