



Date: -14th July, 2017

Corrigendum
For
Tender for
Blood and Fluid Warmer for the Department of
Pediatrics

NIT Issue Date	: 11 th May, 2017
NIT No.	: Admn/Tender/108/2017-AIIMS.JDH
Pre-Bid Meeting	: 23 rd May, 2017 at 03:30 PM
Earlier Last Date of Submission	: 19 th July, 2017 at 03:00 PM
Extended Last Date of Submission	: 01 st August, 2017 at 03:00 PM
Bid opening	: 02 nd August, 2017 at 03:30 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 2**
For
Should be able to maintain or warm fluid/blood at a flow rate of 2.5 L/min
Read
Should be able to maintain or warm fluid/blood at a flow rate of **5-100 mL/min**
- 2. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 3**
For
Should have a digital temperature display of fluid
Read
Should have a digital temperature display or **Numerical display** of fluid
- 3. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 5**
For
Should have a warm water column or heated sleeve up to the patient end to maintain the temperature up to the point of entry into the vein.
Read
Should have a warm water column or heated sleeve **or warming module** up to the patient end to maintain the temperature up to the point of entry into the vein.
- 4. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 8**
For
Should be useful for both in adult and Pediatric patients
Read
Should be useful for both in adult and Pediatric patients **but priming volume should be in a range of 3-5 mL**