



Date: -14<sup>th</sup> July, 2017

Corrigendum  
For  
Tender for  
Patient Warming System for the Department of  
Pediatrics

NIT Issue Date	: 11 <sup>th</sup> May, 2017
NIT No.	: Admn/Tender/110/2017-AIIMS.JDH
Pre-Bid Meeting	: 23 <sup>rd</sup> May, 2017 at 03:30 PM
Earlier Last Date of Submission	: 19 <sup>th</sup> July, 2017 at 03:00 PM
Extended Last Date of Submission	: 01 <sup>st</sup> August, 2017 at 03:00 PM
Bid opening	: 02 <sup>nd</sup> August, 2017 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, Annexure I, Point No. 2**  
**For**  
The convective air patient warming system should have fast warming reaching 38° C within 30 sec.  
**Read**  
The convective air patient warming system should have fast warming reaching 38° C in less than 60 sec.
- 2. Page No. 11, Annexure I, Point No. 4**  
**For**  
The warming system should have an automatic step down facility. After 45 min temperature should come down from high mode to medium mode  
**Read**  
The warming system should have an automatic step down facility.
- 3. Page No. 11, Annexure I, Point No. 9**  
**For**  
Machine should have hour meter to understand total run time  
**Read**  
Machine should have hour meter **or maintain meter** to understand total run time.