



Date: -14th July, 2017

Corrigendum
For
Tender for
Spirometer for the Department of Pediatrics

NIT Issue Date	: 11 th May, 2017
NIT No.	: Admn/Tender/112/2017-AIIMS.JDH
Pre-Bid Meeting	: 23 rd May, 2017 at 03:30 PM
Earlier Last Date of Submission	: 19 th July, 2017 at 03:00 PM
Extended Last Date of Submission	: 01 st August, 2017 at 03:00 PM
Bid opening	: 02 nd August, 2017 at 03:30 P.M

The following revised and additional specification will be added:-

- 1. Page no. 11, Annexure I, Point 1. Operational Requirement, sub-point d.**
For
Feasibility of Bronchial provocation test.
Read
Feasibility of Bronchial provocation test **with dosimeter**
- 2. Page no. 11, Annexure I, Point 2. Technical Specification, sub-point d.**
For
The system should have bi-directional digital turbine flow sensor with following specification
Read
The system should have bi-directional digital turbine **or differential pressure** flow sensor with following specification.
- 3. Page no. 11, Annexure I, Point 2. Technical Specification, sub-point h.**
For
The system should have printer with followings characteristics
- Should have in built printer
- Feasibility of customizable printout options
- Paper should be thermal, 4.4 inch (112 mm) width
Read
The system should have printer with followings characteristics
- Should have in built printer **or have External Printer**
- Feasibility of customizable printout options
- Paper should be thermal, 4.4 inch (112 mm) width