



Date: -14th July, 2017

Corrigendum
For
Tender for
Portable Colour Doppler Ultrasound Machine with 2D
ECHO for the Department of Pediatrics

NIT Issue Date	: 30 th May, 2017
NIT No.	: Admn/Tender/118/2017-AIIMS.JDH
Pre-Bid Meeting	: 09 th June, 2017 at 03:30 PM
Earlier Last Date of Submission	: 27 th July, 2017 at 03:00 PM
Extended Last Date of Submission	: 01 th August, 2017 at 03:00 PM
Bid opening	: 02 nd August, 2017 at 04:00 P.M

The following revised and additional specification will be added:-

- 1. Page No. 11, In Transducers to be supplied as standard, Point No. 01**
For
12-6 MHz broadband linear array with less than 30 mm size for Musculoskeletal, vascular, nerve, superficial, venous, ophthalmic
Read
12-6 ± 1 MHz broadband linear array with less than 30 mm size for Musculoskeletal, vascular, nerve, superficial, venous, ophthalmic
- 2. Page No. 11, In Transducers to be supplied as standard, Point No. 02**
For
8-3 MHz broadband phased array with less than 12 mm size cardiology, abdominal, neonatal head application
Read
8-3 ±1 MHz broadband phased array with less than 12 mm size cardiology, abdominal, neonatal head application
- 3. Page No. 11, In Transducers to be supplied as standard, Point No. 03**
For
8-5 MHz broadband curved array with less than 13 mm size for Abdominal, neonatal, nerve, vascular.
Read
8-5 ±1 MHz broadband curved array with less than 13 mm size for Abdominal, neonatal, nerve, vascular.
- 4. Page No. 11, In Transducers to be supplied as standard, After Point No. 05**
Added Point No. 6:
There should be provision to connect at least 3 transducers with the system