



Date: -31st August, 2017

Corrigendum
For
Tender for
Gynecological Examination Couch for the Department
of Obstetrics & Gynecology

NIT Issue Date	: 05 th June, 2017
NIT No.	: Admn/Tender/120/2017-AIIMS.JDH
Pre-Bid Meeting	: 15 th June, 2017 at 03:15 PM
Earlier Last Date of Submission	01 st August, 2017 at 03:15 PM
Extended Last Date of Submission	: 21 st August, 2017 at 03:00 PM
Bid opening	: 22 nd August, 2017 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, In In Technical Specification, Point No. 10:**
For
Should have minimum patient load capacity of 200 Kg
Read
Should have minimum patient load capacity of 150 Kg
- 2. Page No. 10, In In Technical Specification, Point No. 12, Height:**
For
Height - 550-900 mm
Read
Height - 555-900 mm
- 3. Page No. 10, In In Technical Specification, After Point No. 15:**
Added Point No. 16:
Should conform to water resistant standards of IPX4 or better.