



Date: -31<sup>st</sup> August, 2017

Corrigendum  
For  
Tender for  
Gynecological Examination Couch for the Department  
of Obstetrics & Gynecology

NIT Issue Date	: 05 <sup>th</sup> June, 2017
NIT No.	: Admn/Tender/120/2017-AIIMS.JDH
Pre-Bid Meeting	: 15 <sup>th</sup> June, 2017 at 03:15 PM
Earlier Last Date of Submission	01 <sup>st</sup> August, 2017 at 03:15 PM
Extended Last Date of Submission	: 21 <sup>st</sup> August, 2017 at 03:00 PM
Bid opening	: 22 <sup>nd</sup> August, 2017 at 03:15 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 10, In In Technical Specification, Point No. 10:**  
**For**  
Should have minimum patient load capacity of 200 Kg  
**Read**  
Should have minimum patient load capacity of 150 Kg
- 2. Page No. 10, In In Technical Specification, Point No. 12, Height:**  
**For**  
Height - 550-900 mm  
**Read**  
Height - 555-900 mm
- 3. Page No. 10, In In Technical Specification, After Point No. 15:**  
**Added Point No. 16:**  
Should conform to water resistant standards of IPX4 or better.