



Date: -31st August, 2017

Corrigendum
For
Tender for
Instrument for OPD and Labour Room for the
Department of Obstetrics & Gynecology

NIT Issue Date	: 05 th June, 2017
NIT No.	: Admn/Tender/122/2017-AIIMS.JDH
Pre-Bid Meeting	: 15 th June, 2017 at 03:15 PM
Earlier Last Date of Submission	01 st August, 2017 at 03:15 PM
Extended Last Date of Submission	: 21 st August, 2017 at 03:00 PM
Bid opening	: 22 nd August, 2017 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, In In Technical Specification, After Point No. 25:
Added Point No. 26:**
All instruments should be made of 410/ 420 surgical grade steel.
- 2. Page No. 10, In In Technical Specification, After Point No. 26:
Added Point No. 27:**
Should be having European CE certification or US FDA certification
- 3. Page No. 10, In In Technical Specification, After Point No. 27:
Added Point No. 28:**
Should be able to demonstrate the quoted instruments if asked for.