



Date: - 14th January, 2019

Corrigendum
For
Refrigerated Blood Component Centrifuge for the
Department of Blood Bank

NIT Issue Date	: 01 st November, 2018
NIT No.	: Admn/Tender/124/2018-AIIMS.JDH
Pre-Bid Meeting	: 14 th November, 2018 at 03:00 PM
Earlier Last Date of Submission	: 14 th January, 2019 at 03:00 PM
Extended Last Date of Submission	: 30 th January, 2019 at 03:00 PM
Bid opening	: 31 st January, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page 11, Under sub heading Design and operation, bullet no. 7:

For

“Sixteen 350ml and/or 450ml single, double, triple, quadruple/quintuple blood bags with SAGM bag and empty satellite bags with In Line filter system”.

Read

“Twelve or more 350ml and/or 450ml single, double, triple, quadruple/quintuple blood bags with SAGM bag and empty satellite bags with In Line filter system”.

2. Page 11, Under sub heading Speed and force, sub bullet no. 2:

For

“Maximum RCF (Relative Centrifugal force) for blood bags: 6000g- 65000g”.

Read

“Maximum RCF (Relative Centrifugal force) for blood bags: 5500g-6500g”.

3. Page 12:

For

“Complete with comprehensive set of spare parts and accessories including: Double pan balance, Balancing weights and plates, plastic inserts and spacers and hooks for adjusting to different types and sizes of bag/tubing/filter designs, and a suitable capacity voltage stabilizer and a suitable UPS with maintenance free batteries for minimum one-hour back-up should be supplied free of cost with the system.”

Read

“Complete with comprehensive set of spare parts and accessories including: Double pan balance (electronic), Balancing weights and plates, plastic inserts and spacers and hooks for adjusting to different types and sizes of bag/tubing/filter designs, and a suitable capacity voltage stabilizer should be supplied free of cost with the system”.

4. Page 5, 12th Point: “After Sales Service”:

For

After sales service centre should be available on 24 (hrs.) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 24 hrs to ensure an uptime of minimum 95%, wherever applicable, failing which the necessary penalty measures shall be enforced.

Read

“After Sales Service: After sales service should be available on 24 (hrs.) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 24 hrs to ensure restoration of full functionality or immediate provision of a back-up equipment of same quality within 24 hrs.”

5. Page No. 7 Addendum to be made in “Payment Terms” :

Payment of equipment will be released after 3 months of continued service from the point of completion of IQ-OQ-PQ after installation.