



Date: - 06<sup>th</sup> March, 2019

Corrigendum-II  
For  
Electro Thermal Bipolar Tissue Sealing System for the  
Department of Surgical Gastroenterology

NIT Issue Date	: 01 <sup>st</sup> November, 2018
NIT No.	: Admn/Tender/130/2018-AIIMS.JDH
Pre-Bid Meeting	: 14 <sup>th</sup> November, 2018 at 05:00 PM
Earlier Last Date of Submission	: 28 <sup>th</sup> February, 2019 at 03:00 PM
Extended Last Date of Submission	: 14 <sup>th</sup> March, 2019 at 03:00 PM
Bid opening	: 15 <sup>th</sup> March, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. 1. In corrigendum: Page No. 01, Point 1, Under “Read As”:

**For**

An integrated system with 300W output generator and a single touch screen for Monopolar, Bi-polar and Vessel fusion integrated in one generator or it had only vessel fusion in one generator

**Read**

An integrated system with 300W output generator and a single touch screen/touch button for Monopolar, Bi-polar and Vessel fusion integrated in one generator or it had only vessel fusion in one generator

**NOTE: All the specified open and laparoscopic hand instruments are disposable with similar quantity and dimension as mentioned in corrigendum.**