



Date: - 14th January 2019

Corrigendum
For
Flexible Video Uretero - Renoscope with Chip on the
tip for Urological Procedure with high Definition
System with Recording and Trolley for the
Department of Urology

| | |
|----------------------------------|---|
| NIT Issue Date | : 1 st November, 2018 |
| NIT No. | : Admn/Tender/134/2018-AIIMS.JDH |
| Pre-Bid Meeting | : 14 th November, 2018 at 05:00 PM |
| Earlier Last Date of Submission | : 14 th January, 2019 at 03:00 PM |
| Extended Last Date of Submission | : 30 th January, 2019 at 03:00 PM |
| Bid opening | : 31 st January, 2019 at 03:15 P.M |

The following revised and additional specification will be added:-

- 1. Page No. 10, S. No. 2:**
For
300W Xenon Light Source : 1 No
Read
"Read as 300 Watt xenon or LED light equivalent to 300 watt xenon"
- 2. Page No. 10, S. No. 2. Bullet 2:**
For
Should be 300W Xenon Light Source
Read
"Read as 300 Watt xenon or LED light equivalent to 300 watt xenon"
- 3. Page No. 10, S. No. 3, bullet 1:**
For
26 inch full HD LED monitor with high resolution 1920x1080.
Read
26 inch LCD/LED monitor
- 4. Page No. 10, S. No. 3, bullet 6:**
For
System should offer 2 x DVI-D output, 2 x 3G/HD SDi output
Read
should have necessary Full HD O/P

5. Page No. 10, S. No. 3, bullet 7:

For

Should have 2x DVI – D input and 2 x 3G/HD SDi input

Read

Should have necessary Full HD I/P

6. Page No. 10, S. No. 5 :

For

Flexible Video Uretero-roscope – 3 No's

Read

Flexible Video Uretero-roscope – 2 No's

7. Page No. 11, S. No. 5, bullet 1 :

For

Should be ready to use after only one plug into (existing) HD camera control unit

Read

Should be ready to use after only one plug into HD camera control unit.

8. Page No. 11, S. No. 5, bullet 2 :

For

Scope should have in built light source located at the hand piece of the scope with no external light cable required

Read

Optional if available.

9. Page No. 11, S. No. 5, bullet 3 :

For

Should have the latest state of art CMOS technology for image transmission for better resolution of image

Read

CCD/CMOS

10. Page No. 11, S. No. 5, bullet 16 :

For

Should have a ceramic linear in the distal end of the working channel to protect it from thermal and electro-cautery damage

Read

Optional

11. Page No. 11-12, S. No. 7, bullet 1 :

For

Compatible ureteral access sheath with inner diameter 9.5 Fr. and length of 35 cm. - 5 No's

Read

Compatible ureteral access sheath with inner diameter 9.5 Fr. and length of 35 cm. - 12 No's

12. Page No. 11-12, S. No. 7, bullet 2 :

For

Compatible ureteral access sheath with inner diameter 9.5 Fr. and length of 45 cm. - 5 No's

Read

Compatible ureteral access sheath with inner diameter 9.5 Fr. and length of 45 cm. - 12 No's

13. Page No. 11-12, S. No. 7, bullet 3 :

For

Compatible ureteral access sheath with inner diameter 12 and length of 35 cm. – 5 No's

Read

Compatible ureteral access sheath with inner diameter 12 and length of 35 cm. – 2 No's

14. Page No. 11-12, S. No. 7, bullet 4 :

For

Compatible ureteral access sheath with inner diameter 12 Fr. and length 45 cm. - 5 No's

Read

Compatible ureteral access sheath with inner diameter 12 Fr. and length 45 cm. - 2 No's

15. Page No. 11-12, S. No. 7, bullet 5 :

For

Compatible ureteral access sheath with inner diameter 10.7 Fr. and length of 35 mm. – 5 No's

Read

Deleted

16. Page No. 11-12, S. No. 7, bullet 6 :

For

Compatible ureteral access sheath with inner diameter 10.7 Fr. and length 45 cm. - 5 No's

Read

Deleted

17. There are some additional points that are to be added in the technical specifications:

- a. Full High Definition 3 CCD / 3 Chip CMOS Camera head
- b. It should have 1920x1080 resolution
- c. Camera Head should have 2-4 buttons for features like zoom, white balance, for special imaging mode
- d. System should be US FDA/European CE with 4 digit marking
- e. Demo of quoted model is must