



Date: - 21st June, 2018

Corrigendum
For
Tender for
DEXA Scanner for the Department of Endocrinology
& Metabolism

NIT Issue Date	: 09th April, 2018
NIT No.	: Admn/Tender/14/2018-AIIMS.JDH
Pre-Bid Meeting	: 20th April, 2018 at 03:00 PM
Earlier Last Date of Submission	: 25th June, 2018 at 03:00 PM
Extended Last Date of Submission	: 04th July, 2018 at 03:00 PM
Bid opening	: 05th July, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 11, point 10 (f):**
For
Laser printer
Read
Laser printer with 4 extra cartridge
- 2. Page No. 11, point 11 (n):**
For
(NA)
Read
Trabecular bone scoring software (TBS)
- 3. Page No. 11, point 15th :**
For
20 tonne air conditioning
Read
Minimum 2 tonne air conditioning
- 4. Page No. 11, Point No. 13th :**
For
Online UPS: 2KVA with 30 minutes backup
Read
Online UPS: minimum 2KVA with 30 minutes backup
- 5. Page No. 11**
ADDITION of Points after sub Point 16.4
16.5. Following accessories to be provided
16.5.1. Lead apron (minimum 2)
16.5.2. Computer table & chair

6. Page No. 11, Addition of 17th point after 16th.

17 Bidder to execute site modification works and other miscellaneous works required for successful installation and commissioning of the DEXA scanner