



Date: - 30th July, 2019

Corrigendum
for
CTG Machine for the Department of Obstetrics and
Gynaecology

NIT Issue Date	: 17 th June, 2019
NIT No.	: Admn/Tender/147-2/2018-AIIMS.JDH
Pre-Bid Meeting	: 26 th June, 2019 at 03:00 PM
Earlier Last Date of Submission	: 30 th July, 2019 at 03:00 PM
Extended Last Date of Submission	: 13 th August, 2019 at 03:00 PM
Bid opening	: 14 th August, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 01:

For

7" wide TFT Colour LCD Screen for Fetal Heart Rate, Uterine Contractions & Fetal Movement.

Read

7" wide TFT Colour LCD/LED Screen for Fetal Heart Rate, Uterine Contractions & Fetal Movement.

2. Page No. 10, Point No. 02:

For

LCD Screen should be tiltable, rotatable and can be fixed at any angle from 0-90 degree

Read

LCD/LED screen should be tiltable.

3. Page No. 10, Point No. 7:

For

It should have high sensitivity less than 1 MHz.

Read

It should have high sensitivity **not** less than 1 MHz.

4. Page No. 10, Point No. 19:

For

It should have safety standard certificate like CE or FDA.

Read

It should have safety standard certificate like European CE or US FDA.