



Date: -31st October, 2017

Corrigendum
For
Tender for
-80^o C Deep Freezer (≥ 400 L) for the Department of
Blood Bank

NIT Issue Date	: 15 th September, 2017
NIT No.	: Admn/Tender/162/2017-AIIMS.JDH
Pre-Bid Meeting	: 26 th September, 2017 at 03:00 PM
Earlier Last Date of Submission	: 26 th October, 2017 at 03:00 PM
Last Date of Submission	: 15 th November, 2017 at 03:00 PM
Bid opening	: 16 th November, 2017 at 03:30 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 03:

For

Should have minimum storage volume of 400L or more.

Read

Should have minimum storage volume of 450L or more.

2. Page No. 10, Point No. 19:

For

Should have LED with 0.5^oC display resolution.

Read

Should have LCD/ LED with 0.5^oC display resolution.

3. Page No. 10, Point No. 22:

For

Should have a chart recorder with chart duration of 7 days and chart resolution of 3^o C

Read

Hold over time of a full load of plasma bags should be at least 2 hrs at ambient temperature.

4. Page No. 10, After Point No. 26:

Added Point No. 27:

Equipment should be FDA or European CE approved.

5. Page No. 10, After Point No. 27:

Added Point No. 28:

All relevant Calibration certificates should be submitted at time of installation and IQ-OQ-PQ should be completed by the bidder free of extra charge.