



Date: -22nd September, 2017

Corrigendum
For
Tender for
Image Guided Navigation System for the Department
of ENT

NIT Issue Date	: 15 th September, 2017
NIT No.	: Admn/Tender/165/2017-AIIMS.JDH
Pre-Bid Meeting	: 26 th September, 2017 at 03:30 PM
Last Date of Submission	: 26 th October, 2017 at 03:00 PM
Bid opening	: 27 th October, 2017 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Line No. 50:

For

The system should include a frameless biopsy system with needles (500 disposable needles should be supplied with the system)

Read

The system should include a frameless biopsy system with needles (5 disposable needles should be supplied with the system).

2. Page No. 12, Point No. 18:

For

Should have disposable/reusable wired/wireless electromagnetic standard instrument for patient registration and navigation suitable for 200 surgeries or more. If disposable, a set of disposable instruments set for 200 or more surgeries should be provided.

Read

Should have disposable/reusable wired/wireless electromagnetic standard instrument for patient registration and navigation suitable for multiple surgeries. If disposable, minimum 10 sets of disposable instruments should be provided with the unit.